



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96446

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	07/05/2018	15:15
2	CAL2	1	Water			21.2	7.00	07/05/2018	15:16
3	CAL3	1	Water			21.2	10.03	07/05/2018	15:17
4	ICV	1	Water			21.3	7.01	07/05/2018	15:19
5	CCV1	1	Water			21.2	2.01	07/05/2018	15:20
6	J3878-01	1	Water			20.7	6.01	07/05/2018	15:21
7	J3878-01DUP	1	Water			20.8	6.02	07/05/2018	15:22
8	J3878-03	1	Water			20.6	6.45	07/05/2018	15:25
9	J3878-04	1	Water			21.4	6.54	07/05/2018	15:26
10	CCV2	1	Water			21.1	12.02	07/05/2018	15:29

WorkList Name : PH J3878

WorkList ID : 114142

Date : 7/5/2018 3:08:01 PM

WORKLIST(Hardcopy Internal Chain)

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2018	Water	J3878-01-Cy	pH	Cool 4 deg C	LOCK01	I33	MW-1	07/03/2018	9045C
07/03/2018	Water	J3878-03 F	pH	Cool 4 deg C	LOCK01	I33	MW-3	07/03/2018	9045C
07/03/2018	Water	J3878-04 P	pH	Cool 4 deg C	LOCK01	I33	MW-4	07/03/2018	9045C

Date/Time

Received by:

Relinquished by:

07-05-18 3:13 PM

JP

CE

Date/Time

Received by:

Relinquished by:

07-05-18 3:10 PM

CE

JP