



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96453

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/05/2018	16:45
2	CAL2	1	Water			21.6	7.00	07/05/2018	16:47
3	CAL3	1	Water			21.8	10.03	07/05/2018	16:49
4	ICV	1	Water			21.6	7.00	07/05/2018	16:51
5	CCV1	1	Water			21.7	2.00	07/05/2018	16:53
6	J3877-02	1	Solid	20.02	20	24.2	9.11	07/05/2018	16:55
7	J3877-02DUP	1	Solid	20.04	20	24.2	9.12	07/05/2018	16:57
8	J3877-03	1	Solid	20.03	20	24.1	6.29	07/05/2018	16:59
9	J3887-04	1	Solid	20.02	20	24.2	9.22	07/05/2018	17:01
10	J3887-08	1	Solid	20.04	20	24.4	8.65	07/05/2018	17:03
11	CCV2	1	Water			21.7	12.00	07/05/2018	17:05

WORKLIST(Hardcopy Internal Chain)

LB 96453

WorkList Name : corrosivity070518

WorkList ID : 114149

Date : 7/5/2018 4:32:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/05/2018	Solid	J3877-02	Corrosivity	Cool 4 deg C	PSEG01	I42	NS-DRUM-2-STONE	07/05/2018	9045C
07/05/2018	Solid	J3877-03	Corrosivity	Cool 4 deg C	PSEG01		NS-DRUM-3-OIL	07/05/2018	9045C
07/05/2018	Solid	J3887-04	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-8-S	07/05/2018	9045C
07/05/2018	Solid	J3887-08	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-8-D	07/05/2018	9045C

Date/Time 07/05/18 4:40 PM

Received by: SP

Relinquished by: SP

Date/Time 07/05/18 5:36 PM

Received by: SP

Relinquished by: SP