



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96454

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/05/2018	16:45
2	CAL2	1	Water			21.6	7.00	07/05/2018	16:47
3	CAL3	1	Water			21.8	10.03	07/05/2018	16:49
4	ICV	1	Water			21.6	7.00	07/05/2018	16:51
5	CCV1	1	Water			21.7	2.00	07/05/2018	17:07
6	J3877-04	1	Water			23.7	6.01	07/05/2018	17:09
7	J3877-04DUP	1	Water			23.8	6.02	07/05/2018	17:11
8	J3877-05	1	Water			24.0	5.95	07/05/2018	17:13
9	CCV2	1	Water			21.6	12.00	07/05/2018	17:15

WORKLIST(Hardcopy Internal Chain)

LB96454

WorkList Name : corrosivity070518-1

WorkList ID : 114150

Date : 7/5/2018 4:32:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/05/2018	Water	J3877-04	Corrosivity	Cool 4 deg C	PSEG01	I42	NS-DRUM-1-WATER	07/05/2018	9040C
07/05/2018	Water	J3877-05	Corrosivity	Cool 4 deg C	PSEG01	I42	NS-DRUM-4-WATER	07/05/2018	9040C

Date/Time 07/05/18 4:40PM

Received by: SP

Relinquished by: CP

Date/Time 07/05/18 5:30PM

Received by: CP

Relinquished by: SP