



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96467

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP65865
Turbidity std, 40NTU	WP65860
turbidity std 5 NTU	WP65863
turbidity std 20 NTU	WP65861
Turbidity std 1 NTU	WP65864
turbidity std 0 NTU	WP65866
turbidity std 10 NTU	WP65862
10 NTU Standard 500 ml	W2242

Intercept: -0.0464

Slope: 0.9758

Regression: 0.9995

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.241	07/05/2018	12:50
2	CAL2	1	1	0.992	07/05/2018	12:54
3	CAL3	5	1	4.890	07/05/2018	12:59
4	CAL4	10	1	9.760	07/05/2018	13:04
5	CAL5	20	1	18.600	07/05/2018	13:08
6	CAL6	40	1	39.400	07/05/2018	13:12



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96467

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.320	9.599	07/05/2018	13:16
2	ICB		1	0.236	0.289	07/05/2018	13:20
3	CCV1	10	1	9.420	9.701	07/05/2018	13:25
4	CCB1		1	0.235	0.288	07/05/2018	13:31
5	LB96467BL		1	0.238	0.291	07/05/2018	13:35
6	J3878-01		1	0.893	0.963	07/05/2018	13:39
7	J3878-01DUP		1	0.890	0.960	07/05/2018	13:43
8	J3878-03		1	0.595	0.657	07/05/2018	13:47
9	J3878-04		1	0.645	0.709	07/05/2018	13:50
10	CCV2	10	1	9.560	9.845	07/05/2018	13:54
11	CCB2		1	0.208	0.261	07/05/2018	13:58

CB96467

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity070518

WorkList ID : 114136

Date : 7/5/2018 12:25:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/05/2018	Water	J3878-01	Turbidity	Cool 4 deg C	LOCK01	I33	MW-1	07/03/2018	SM2130 B
07/05/2018	Water	J3878-03	Turbidity	Cool 4 deg C	LOCK01	I33	MW-3	07/03/2018	SM2130 B
07/05/2018	Water	J3878-04	Turbidity	Cool 4 deg C	LOCK01	I33	MW-4	07/03/2018	SM2130 B

Date/Time 07-05-18 12:35
 Received by: AE
 Relinquished by: CR

Date/Time 7-5-18 14:10
 Received by: Q
 Relinquished by: AE