



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96471

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	07/06/2018	09:40
2	J3877-04		1	7	>100.0	>212.0	>212.0	07/06/2018	10:05
3	J3877-04DUP		1	7	>100.0	>212.0	>212.0	07/06/2018	10:40
4	J3877-05		1	8	>100.0	>212.0	>212.0	07/06/2018	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96471

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-070618

WorkList ID : 114176

Date : 7/6/2018 8:59:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2018	Water	J3877-04 B	Flash Point	Cool 4 deg C	PSEG01	I42	NS-DRUM-1-WATER	07/05/2018	1010A
07/12/2018	Water	J3877-05 B	Flash Point	Cool 4 deg C	PSEG01	I42	NS-DRUM-4-WATER	07/05/2018	1010A

Date/Time

Received by:

Relinquished by:

07/06/18 9:10 AM

KS

CP

Date/Time

Received by:

Relinquished by:

07/06/18 11:15 AM

CP

KS