



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96474

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP65874
Turbidity std, 80NTU	WP65867
Turbidity std, 40NTU	WP65868
turbidity std 5 NTU	WP65871
turbidity std 20 NTU	WP65869
Turbidity std 1 NTU	WP65872
turbidity std 0 NTU	WP65873
turbidity std 10 NTU	WP65870
10 NTU Standard 500 ml	W2242

Intercept: -0.1455

Slope: 0.9972

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.209	07/06/2018	13:07
2	CAL2	1	1	0.997	07/06/2018	13:10
3	CAL3	5	1	4.980	07/06/2018	13:14
4	CAL4	10	1	9.960	07/06/2018	13:21
5	CAL5	20	1	18.900	07/06/2018	13:23
6	CAL6	40	1	39.600	07/06/2018	13:26
7	CAL7	80	1	79.900	07/06/2018	13:31



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96474

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.420	9.592	07/06/2018	13:36
2	ICB		1	0.210	0.356	07/06/2018	13:40
3	CCV1	10	1	9.780	9.953	07/06/2018	13:43
4	CCB1		1	0.241	0.388	07/06/2018	13:46
5	LB96474BL		1	0.216	0.363	07/06/2018	13:50
6	J3878-02		1	76.500	76.861	07/06/2018	13:53
7	J3878-02DUP		1	75.200	75.557	07/06/2018	13:56
8	CCV2	10	1	9.800	9.973	07/06/2018	13:59
9	CCB2		1	0.229	0.376	07/06/2018	14:03

0896474

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tur070618

WorkList ID : 114187

Date : 7/6/2018 12:35:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/05/2018	Water	J3878-02	Turbidity	Cool 4 deg C	LOCK01	I33	MW-2	07/03/2018	SM2130 B

Date/Time 7-6-18 12:40
Received by: AK
Relinquished by: OP

Date/Time 7-6-18 14:08
Received by: OP
Relinquished by: AK