

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB96479

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP65884
calibration std. phosphate 0.5 ppm	WP65883
calibration std. phosphate 0.3 ppm	WP65882
calibration std. 0 ppm	WP65879
phosphate ICV std.	WP65886
phosphate CCV std.	WP65885
calibration std. phosphate 0.1 ppm	WP65881
calibration std. phosphate 0.05 ppm	WP65880
Combined reagent	WP65888
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0011Slope: 0.6511Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/06/2018	11:30
2	CAL2	0.05	1	50	50	0.033	07/06/2018	11:31
3	CAL3	0.10	1	50	50	0.065	07/06/2018	11:31
4	CAL4	0.30	1	50	50	0.186	07/06/2018	11:32
5	CAL5	0.50	1	50	50	0.329	07/06/2018	11:32
6	CAL6	1.00	1	50	50	0.650	07/06/2018	11:33



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB96479

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	07/06/2018	11:33
2	ICB		1	50	50	0.000	0.002	07/06/2018	11:34
3	CCV1	0.50	1	50	50	0.328	0.505	07/06/2018	11:35
4	CCB1		1	50	50	0.000	0.002	07/06/2018	11:35
5	PB110907BL		1	50	50	0.000	0.002	07/06/2018	11:35
6	PB110907BS	0.50	1	50	50	0.327	0.504	07/06/2018	11:36
7	J3838-01		1	50	50	0.982	1.510	07/06/2018	11:36
8	J3838-01DUP		1	50	50	0.997	1.533	07/06/2018	11:37
9	J3838-01MS	0.50	1	50	50	1.026	1.577	07/06/2018	11:38
10	J3838-01MSD	0.50	1	50	50	1.027	1.579	07/06/2018	11:38
11	J3838-01		5	50	50	0.466	0.717	07/06/2018	11:50
12	J3838-01DUP		5	50	50	0.462	0.711	07/06/2018	11:50
13	J3838-01MS	0.50	5	50	50	0.525	0.808	07/06/2018	11:52
14	J3838-01MSD	0.50	5	50	50	0.524	0.806	07/06/2018	11:52
15	CCV2	0.50	1	50	50	0.325	0.501	07/06/2018	11:53
16	CCB2		1	50	50	0.000	0.002	07/06/2018	11:53