



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96499

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	07/06/2018	15:20
2	CAL2	1	Water			21.4	7.01	07/06/2018	15:21
3	CAL3	1	Water			21.4	10.03	07/06/2018	15:22
4	ICV	1	Water			21.3	6.99	07/06/2018	15:24
5	CCV1	1	Water			21.3	2.01	07/06/2018	15:25
6	J3878-02	1	Water			22.1	6.49	07/06/2018	15:26
7	J3878-02DUP	1	Water			22.2	6.50	07/06/2018	15:29
8	CCV2	1	Water			21.4	12.01	07/06/2018	15:30

WORKLIST(Hardcopy Internal Chain)

LB 96499

WorkList Name : pH J3878-1

WorkList ID : 114194

Date : 7/6/2018 3:05:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2018	Water	J3878-02	pH	Cool 4 deg C	LOCK01	I33	MW-2	07/03/2018	SM 4500-PH B

Date/Time 07/06/18 3:10
 Received by: SP
 Relinquished by: CP

Date/Time 07/06/18 4:35 PM
 Received by: CP
 Relinquished by: SP