


7471B

LB 96502
INSTRUMENT ID: CV 2

Linear

 μ Abs.:

47064



A= 0.0000e+000

B= 2.1286e-004

C= 2.4166e-002

Rho= 0.9999425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	30	0.000	30				
0.2	0.200	0.197	-0.003	812	0.0 %	812				
2.5	2.500	2.466	-0.034	11471	0.0 %	11471				
5.0	5.000	5.031	0.031	23523	0.0 %	23523				
7.5	7.500	7.433	-0.067	34805	0.0 %	34805				
10.0	10.000	10.042	0.042	47064	0.0 %	47064				

LB96502 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jul 2018 09:00:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	30	-	0.00007471B	PPB		09 Jul 2018 09:05:20	S	Std
0.2 - 1	50.2	812	-	0.20007471B	PPB		09 Jul 2018 09:07:38	S	Std
2.5 - 1	52.5	11471	-	2.50007471B	PPB		09 Jul 2018 09:09:55	S	Std
5.0 - 1	55	23523	-	5.00007471B	PPB		09 Jul 2018 09:12:13	S	Std
7.5 - 1	57.5	34805	-	7.50007471B	PPB		09 Jul 2018 09:14:30	S	Std
10.0 - 1	510	47064	-	10.00007471B	PPB		09 Jul 2018 09:19:44	S	Std
ICV38 - 1	ICV38	18435	3.9483	-7471B	PPB		09 Jul 2018 09:23:25	U	SMPL
ICB38 - 1	ICB38	-109	0.0010	-7471B	PPB		09 Jul 2018 09:25:40	U	SMPL
CCV24 - 1	CCV24	23800	5.0903	-7471B	PPB		09 Jul 2018 09:27:56	U	SMPL
CCB24 - 1	CCB24	-129	-0.0033	-7471B	PPB		09 Jul 2018 09:30:12	U	SMPL
CRA - 1	CRA	813	0.1972	-7471B	PPB		09 Jul 2018 09:34:17	U	SMPL
HighStd - 1	HighStd	46945	10.0170	-7471B	PPB		09 Jul 2018 09:36:33	U	SMPL
ChkStd - 1	ChkStd	33183	7.0876	-7471B	PPB		09 Jul 2018 09:38:49	U	SMPL
PB110876BL - 1	PBS	-141	-0.0058	-7471B	PPB		09 Jul 2018 09:41:07	U	SMPL
PB110876BS - 1	LCSS	17409	3.7299	-7471B	PPB		09 Jul 2018 09:43:24	U	SMPL
J3845-01 - 1	GF-01	1700	0.3860	-7471B	PPB		09 Jul 2018 09:45:41	U	SMPL
J3845-02 - 1	GF-02	1577	0.3599	-7471B	PPB		09 Jul 2018 09:47:58	U	SMPL
J3845-02DUP - 1	GF-02DUP	1429	0.3283	-7471B	PPB		09 Jul 2018 09:50:15	U	SMPL
J3845-03 - 1	J3845-02MS	18582	3.9796	-7471B	PPB		09 Jul 2018 09:52:33	U	SMPL
J3845-04 - 1	J3845-02MSD	18650	3.9941	-7471B	PPB		09 Jul 2018 09:54:49	U	SMPL
CCV25 - 1	CCV25	23539	5.0347	-7471B	PPB		09 Jul 2018 09:57:06	U	SMPL
CCB25 - 1	CCB25	-139	-0.0054	-7471B	PPB		09 Jul 2018 09:59:22	U	SMPL
J3877-01 - 1	NS-DRUM-2-STONE	1391	0.3203	-7471B	PPB		09 Jul 2018 10:01:39	U	SMPL
J3881-01 - 1	SURCHARGE-SP-21	16492	3.5347	-7471B	PPB		09 Jul 2018 10:03:55	U	SMPL
J3881-03 - 1	SURCHARGE-SP-22	18008	3.8574	-7471B	PPB		09 Jul 2018 10:06:11	U	SMPL
J3881-05 - 1	SURCHARGE-SP-23	18066	3.8697	-7471B	PPB		09 Jul 2018 10:08:27	U	SMPL
J3881-07 - 1	SURCHARGE-SP-24	19431	4.1603	-7471B	PPB		09 Jul 2018 10:10:44	U	SMPL
J3881-09 - 1	SURCHARGE-SP-25	22024	4.7123	-7471B	PPB		09 Jul 2018 10:13:01	U	SMPL
J3886-01 - 1	EPE-108-6A(55-57)	936	0.2234	-7471B	PPB		09 Jul 2018 10:15:19	U	SMPL
J3886-02 - 1	EPE-108-6A(60-62)	508	0.1323	-7471B	PPB		09 Jul 2018 10:17:37	U	SMPL
J3887-03 - 1	TP-8-S	1148	0.2685	-7471B	PPB		09 Jul 2018 10:19:55	U	SMPL
J3887-07 - 1	TP-8-D	207	0.0682	-7471B	PPB		09 Jul 2018 10:22:11	U	SMPL
CCV26 - 1	CCV26	23730	5.0754	-7471B	PPB		09 Jul 2018 10:24:27	U	SMPL
CCB26 - 1	CCB26	-153	-0.0084	-7471B	PPB		09 Jul 2018 10:26:43	U	SMPL
PB110877BL - 1	PBS	-17	0.0205	-7471B	PPB		09 Jul 2018 10:29:00	U	SMPL
PB110877BS - 1	LCSS	16642	3.5666	-7471B	PPB		09 Jul 2018 10:31:16	U	SMPL
J3879-01 - 1	SURCHARGE-SP-1	13535	2.9053	-7471B	PPB		09 Jul 2018 10:33:32	U	SMPL
J3879-01DUP - 1	SURCHARGE-SP-1DUP	15474	3.3180	-7471B	PPB		09 Jul 2018 10:35:49	U	SMPL
J3879-01MS - 1	SURCHARGE-SP-1MS	30651	6.5486	-7471B	PPB		09 Jul 2018 10:38:05	U	SMPL
J3879-01MSD - 1	SURCHARGE-SP-1MSD	29885	6.3856	-7471B	PPB		09 Jul 2018 10:40:22	U	SMPL
J3879-03 - 1	SURCHARGE-SP-2	14024	3.0094	-7471B	PPB		09 Jul 2018 10:42:41	U	SMPL
J3879-05 - 1	SURCHARGE-SP-3	14028	3.0102	-7471B	PPB		09 Jul 2018 10:44:59	U	SMPL
J3879-07 - 1	SURCHARGE-SP-4	10019	2.1568	-7471B	PPB		09 Jul 2018 10:47:18	U	SMPL
J3879-09 - 1	SURCHARGE-SP-5	13885	2.9798	-7471B	PPB		09 Jul 2018 10:49:34	U	SMPL
CCV27 - 1	CCV27	23947	5.1216	-7471B	PPB		09 Jul 2018 10:51:50	U	SMPL
CCB27 - 1	CCB27	-150	-0.0078	-7471B	PPB		09 Jul 2018 10:54:07	U	SMPL
J3879-11 - 1	SURCHARGE-SP-6	12413	2.6664	-7471B	PPB		09 Jul 2018 10:56:24	U	SMPL
J3879-13 - 1	SURCHARGE-SP-7	16796	3.5994	-7471B	PPB		09 Jul 2018 10:58:40	U	SMPL
J3879-15 - 1	SURCHARGE-SP-8	16555	3.5481	-7471B	PPB		09 Jul 2018 11:00:56	U	SMPL
J3879-17 - 1	SURCHARGE-SP-9	18458	3.9532	-7471B	PPB		09 Jul 2018 11:03:13	U	SMPL
J3879-19 - 1	SURCHARGE-SP-10	16765	3.5928	-7471B	PPB		09 Jul 2018 11:05:30	U	SMPL
J3880-01 - 1	SURCHARGE-SP-11	16656	3.5696	-7471B	PPB		09 Jul 2018 11:07:46	U	SMPL
J3880-03 - 1	SURCHARGE-SP-12	15476	3.3184	-7471B	PPB		09 Jul 2018 11:10:04	U	SMPL
J3880-05 - 1	SURCHARGE-SP-13	8943	1.9278	-7471B	PPB		09 Jul 2018 11:12:21	U	SMPL
J3880-07 - 1	SURCHARGE-SP-14	10734	2.3090	-7471B	PPB		09 Jul 2018 11:14:39	U	SMPL
J3880-09 - 1	SURCHARGE-SP-15	11325	2.4348	-7471B	PPB		09 Jul 2018 11:16:55	U	SMPL
CCV28 - 1	CCV28	23683	5.0654	-7471B	PPB		09 Jul 2018 11:19:12	U	SMPL
CCB28 - 1	CCB28	-148	-0.0073	-7471B	PPB		09 Jul 2018 11:21:28	U	SMPL
J3880-11 - 1	SURCHARGE-SP-16	11580	2.4891	-7471B	PPB		09 Jul 2018 11:23:46	U	SMPL
J3880-13 - 1	SURCHARGE-SP-17	10707	2.3033	-7471B	PPB		09 Jul 2018 11:26:02	U	SMPL
J3880-15 - 1	SURCHARGE-SP-18	10318	2.2205	-7471B	PPB		09 Jul 2018 11:28:18	U	SMPL
J3880-17 - 1	SURCHARGE-SP-19	11077	2.3820	-7471B	PPB		09 Jul 2018 11:30:35	U	SMPL
J3880-19 - 1	SURCHARGE-SP-20	15801	3.3876	-7471B	PPB		09 Jul 2018 11:32:51	U	SMPL
J3845-02LX5 - 1		154	0.0569	-7471B	PPB		09 Jul 2018 11:35:08	U	SMPL
J3845-02A - 1		19133	4.0969	-7471B	PPB		09 Jul 2018 11:37:26	U	SMPL
J3879-01LX5 - 1		2469	0.5497	-7471B	PPB		09 Jul 2018 11:39:43	U	SMPL
J3879-01A - 1		32508	6.9439	-7471B	PPB		09 Jul 2018 11:42:02	U	SMPL
CCV29 - 1	CCV29	24091	5.1522	-7471B	PPB		09 Jul 2018 11:44:19	U	SMPL
CCB29 - 1	CCB29	-166	-0.0112	-7471B	PPB		09 Jul 2018 11:46:37	U	SMPL