

Analytical Summary ReportAnalysis Method: SM5220 DANALYST: AMANDEEPParameter: CODSUPERVISOR REVIEW BY: apatelRun Number: LB96505

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP65785
COD calibration std. 100 ppm	WP65783
COD calibration std. 10 ppm	WP65781
COD calibration std. 150 ppm	WP65784
COD calibration std. 50 ppm	WP65782
COD calibration std. 0 ppm	WP65778
COD CCV std, 50ppm	WP65786
COD Digestion Vials Low Level 0-150Mg/L	W2375

Temp In (C): <u>150</u>	Date In: <u>07/09/2018</u>	Time In: <u>09:22</u>
Temp Out (C): <u>150</u>	Date Out: <u>07/09/2018</u>	Time Out: <u>11:24</u>

Intercept: -0.2985Slope: 0.9984Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/03/2018	13:00
2	CAL2	10	1	Water	10.000	07/03/2018	13:00
3	CAL3	50	1	Water	49.000	07/03/2018	13:01
4	CAL4	100	1	Water	99.000	07/03/2018	13:01
5	CAL5	150	1	Water	150.000	07/03/2018	13:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: AMANDEEP

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB96505

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.379	07/09/2018	11:56
2	ICB				1	Water	0.000	0.299	07/09/2018	11:56
3	CCV1	50			1	Water	51.000	51.381	07/09/2018	11:57
4	CCB1				1	Water	0.000	0.299	07/09/2018	11:57
5	LB96505BL				1	Water	0.000	0.299	07/09/2018	11:58
6	LB96505BS	50			1	Water	50.000	50.379	07/09/2018	11:58
7	J3866-03				1	Water	103.000	103.464	07/09/2018	11:59
8	J3866-03DUP				1	Water	107.000	107.470	07/09/2018	11:59
9	J3866-03MS	50			1	Water	145.000	145.531	07/09/2018	12:00
10	J3866-03MSD	50			1	Water	149.000	149.538	07/09/2018	12:01
11	CCV2	50			1	Water	50.000	50.379	07/09/2018	12:02
12	CCB2				1	Water	0.000	0.299	07/09/2018	12:02

LB 96505 WORKLIST(Hardcopy Internal Chain)

WorkList Name : cod070918

WorkList ID : 114217

Date : 7/9/2018 8:56:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/06/2018	Water	J3866-03 E	COD	Conc H2SO4 to	PSEG01	D62	AG-OILY-WATER	07/03/2018	SM5220 D

Date/Time 7-9-18 9:00
 Received by: AR
 Relinquished by: CP

Date/Time 7-9-18 10:00
 Received by: CP
 Relinquished by: AR