


245.1

LB 96507
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35387



A= 0.0000e+000

B= 2.8432e-004

C= -1.2993e-002

Rho= 0.9999261

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	-54	0.000	-54				
0.05	0.050	0.046	-0.004	207	0.0 %	207				
0.2	0.200	0.220	0.020	820	0.0 %	820				
2.5	2.500	2.571	0.071	9087	0.0 %	9087				
5.0	5.000	4.951	-0.049	17460	0.0 %	17460				
7.5	7.500	7.442	-0.058	26221	0.0 %	26221				
10.0	10.000	10.048	0.048	35387	0.0 %	35387				

LB96507 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Jul 2018 13:40:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-54	-	0.0000	245.1	PPB	09 Jul 2018 13:42:19	S	Std
0.05 - 1	50.05	207	-	0.0500	245.1	PPB	09 Jul 2018 13:44:27	S	Std
0.2 - 1	50.2	820	-	0.2000	245.1	PPB	09 Jul 2018 13:46:35	S	Std
2.5 - 1	52.5	9087	-	2.5000	245.1	PPB	09 Jul 2018 13:48:44	S	Std
5.0 - 1	55	17460	-	5.0000	245.1	PPB	09 Jul 2018 13:50:52	S	Std
7.5 - 1	57.5	26221	-	7.5000	245.1	PPB	09 Jul 2018 13:52:59	S	Std
10.0 - 1	510	35387	-	10.0000	245.1	PPB	09 Jul 2018 13:57:51	S	Std
ICV39 - 1	ICV39	13729	3.8904	-245.1		PPB	09 Jul 2018 14:04:12	U	SMPL
ICB39 - 1	ICB39	-212	-0.0733	-245.1		PPB	09 Jul 2018 14:06:18	U	SMPL
CCV30 - 1	CCV30	17943	5.0886	-245.1		PPB	09 Jul 2018 14:08:24	U	SMPL
CCB30 - 1	CCB30	-235	-0.0798	-245.1		PPB	09 Jul 2018 14:10:30	U	SMPL
CRA - 1	CRA	238	0.0547	-245.1		PPB	09 Jul 2018 14:15:00	U	SMPL
HighStd - 1	HighStd	35164	9.9849	-245.1		PPB	09 Jul 2018 14:17:06	U	SMPL
ChkStd - 1	ChkStd	24977	7.0885	-245.1		PPB	09 Jul 2018 14:19:13	U	SMPL
PB110847BL - 1	PBW	-269	-0.0895	-245.1		PPB	09 Jul 2018 14:21:22	U	SMPL
PB110847BS - 1	LCSW	13414	3.8009	-245.1		PPB	09 Jul 2018 14:23:31	U	SMPL
J3851-01 - 1	070218-DBRI-F-D-B	-251	-0.0844	-245.1		PPB	09 Jul 2018 14:28:03	U	SMPL
J3851-02 - 1	070218-DBRI-F-D-S	-239	-0.0809	-245.1		PPB	09 Jul 2018 14:30:10	U	SMPL
J3851-03 - 1	070218-DBRI-F-U-B	-239	-0.0809	-245.1		PPB	09 Jul 2018 14:32:18	U	SMPL
J3851-04 - 1	070218-DBRI-F-U-S	-237	-0.0804	-245.1		PPB	09 Jul 2018 14:34:27	U	SMPL
J3851-05 - 1	070218-TIPC-F-D-B	-250	-0.0841	-245.1		PPB	09 Jul 2018 14:36:34	U	SMPL
CCV31 - 1	CCV31	17624	4.9979	-245.1		PPB	09 Jul 2018 14:38:40	U	SMPL
CCB31 - 1	CCB31	-238	-0.0807	-245.1		PPB	09 Jul 2018 14:40:47	U	SMPL
J3851-06 - 1	070218-TIPC-F-D-S	-251	-0.0844	-245.1		PPB	09 Jul 2018 14:42:55	U	SMPL
J3851-07 - 1	070218-TIPC-F-U-B	-241	-0.0815	-245.1		PPB	09 Jul 2018 14:45:02	U	SMPL
J3851-08 - 1	070218-TIPC-F-U-S	-251	-0.0844	-245.1		PPB	09 Jul 2018 14:47:09	U	SMPL
J3857-01 - 1	COMP	-132	-0.0505	-245.1		PPB	09 Jul 2018 14:49:16	U	SMPL
J3857-01DUP - 1	COMP DUP	-143	-0.0537	-245.1		PPB	09 Jul 2018 14:51:23	U	SMPL
J3857-01MS - 1	COMPMS	13438	3.8077	-245.1		PPB	09 Jul 2018 14:53:30	U	SMPL
J3857-01MSD - 1	COMPMSD	13570	3.8452	-245.1		PPB	09 Jul 2018 14:55:38	U	SMPL
J3857-01LX5 - 1		-219	-0.0753	-245.1		PPB	09 Jul 2018 14:57:47	U	SMPL
J3857-01A - 1		14018	3.9726	-245.1		PPB	09 Jul 2018 14:59:55	U	SMPL
CCV32 - 1	CCV32	17874	5.0690	-245.1		PPB	09 Jul 2018 15:02:02	U	SMPL
CCB32 - 1	CCB32	-227	-0.0775	-245.1		PPB	09 Jul 2018 15:04:10	U	SMPL