



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96514

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/09/2018	15:40
2	CAL2	1	Water			21.9	7.00	07/09/2018	15:42
3	CAL3	1	Water			21.7	10.03	07/09/2018	15:44
4	ICV	1	Water			21.7	7.00	07/09/2018	15:46
5	CCV1	1	Water			21.8	2.00	07/09/2018	15:48
6	J3911-01	1	Water			23.3	7.29	07/09/2018	15:50
7	J3911-01DUP	1	Water			23.4	7.30	07/09/2018	15:52
8	J3911-03	1	Water			23.2	6.90	07/09/2018	15:54
9	CCV2	1	Water			21.8	12.00	07/09/2018	15:56

WORKLIST(Hardcopy Internal Chain)

LB 96514

WorkList Name : pH070918

WorkList ID : 114239

Date : 7/9/2018 3:14:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2018	Water	J3911-01	pH	Cool 4 deg C	PSEG01	K12	RB47196RT	07/09/2018	9040C
07/09/2018	Water	J3911-03	pH	Cool 4 deg C	PSEG01	K12	VB21029	07/09/2018	9040C

Date/Time 07/09/18 3:30
 Received by: SP
 Relinquished by: CP

Date/Time 07/09/18 4:10
 Received by: CP
 Relinquished by: SP