



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96516

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/09/2018	15:40
2	CAL2	1	Water			21.9	7.00	07/09/2018	15:42
3	CAL3	1	Water			21.7	10.03	07/09/2018	15:44
4	ICV	1	Water			21.7	7.00	07/09/2018	15:46
5	CCV1	1	Water			21.8	2.00	07/09/2018	16:30
6	J3827-02	1	Solid	20.02	20	24.2	8.52	07/09/2018	16:32
7	J3827-02DUP	1	Solid	20.03	20	24.3	8.53	07/09/2018	16:34
8	J3908-04	1	Solid	20.02	20	23.6	9.33	07/09/2018	16:36
9	J3914-02	1	Solid	20.04	20	24.1	8.03	07/09/2018	16:38
10	J3914-04	1	Solid	20.02	20	24.4	8.05	07/09/2018	16:40
11	J3914-06	1	Solid	20.04	20	24.3	8.28	07/09/2018	16:42
12	J3914-08	1	Solid	20.02	20	23.6	8.56	07/09/2018	16:44
13	J3914-10	1	Solid	20.02	20	23.8	8.26	07/09/2018	16:46
14	J3914-12	1	Solid	20.03	20	23.6	8.07	07/09/2018	16:48
15	CCV2	1	Water			21.7	12.00	07/09/2018	16:50

WORKLIST(Hardcopy Internal Chain)

LB 96516

WorkList Name : corrosivity070918

WorkList ID : 114237

Date : 7/9/2018 3:14:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2018	Solid	J3827-02	Corrosivity	Cool 4 deg C	PSEG05	K12	OR-03-070918	07/09/2018	9045C
07/09/2018	Solid	J3908-04	Corrosivity	Cool 4 deg C	PSEG01	K11	TP-10-S	07/09/2018	9045C
07/09/2018	Solid	J3914-02	Corrosivity	Cool 4 deg C	PSEG01	K13	B1	07/09/2018	9045C
07/09/2018	Solid	J3914-04	Corrosivity	Cool 4 deg C	PSEG01	K13	B2	07/09/2018	9045C
07/09/2018	Solid	J3914-06	Corrosivity	Cool 4 deg C	PSEG01	K13	B3	07/09/2018	9045C
07/09/2018	Solid	J3914-08	Corrosivity	Cool 4 deg C	PSEG01	K13	NEMH	07/09/2018	9045C
07/09/2018	Solid	J3914-10	Corrosivity	Cool 4 deg C	PSEG01	K13	SEMH	07/09/2018	9045C
07/09/2018	Solid	J3914-12	Corrosivity	Cool 4 deg C	PSEG01	K13	SMH	07/09/2018	9045C

Date/Time 07/09/18 4:15
 Received by: SP
 Relinquished by: SP

Date/Time 07/09/18 5:00
 Received by: SP
 Relinquished by: SP