


7471B

L396523
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45284



A= 0.0000e+000

B= 2.2228e-004

C= -9.7752e-002

Rho= 0.9999266

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.050	-0.050	217	0.000	217				
0.2	0.200	0.208	0.008	1376	0.0 %	1376				
2.5	2.500	2.504	0.004	11705	0.0 %	11705				
5.0	5.000	5.090	0.090	23337	0.0 %	23337				
7.5	7.500	7.480	-0.020	34091	0.0 %	34091				
10.0	10.000	9.968	-0.032	45284	0.0 %	45284				

LB96523 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Jul 2018 08:16:41

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	217	-	0.00007471B	PPB	10 Jul 2018 08:21:57	S	Std
0.2 - 1	50.2	1376	-	0.20007471B	PPB	10 Jul 2018 08:24:14	S	Std
2.5 - 1	52.5	11705	-	2.50007471B	PPB	10 Jul 2018 08:26:32	S	Std
5.0 - 1	55	23337	-	5.00007471B	PPB	10 Jul 2018 08:28:50	S	Std
7.5 - 1	57.5	34091	-	7.50007471B	PPB	10 Jul 2018 08:31:07	S	Std
10.0 - 1	510	45284	-	10.00007471B	PPB	10 Jul 2018 08:36:02	S	Std
ICV40 - 1	ICV40	18726	4.0646	-7471B	PPB	10 Jul 2018 08:38:50	U	SMPL
ICB40 - 1	ICB40	137	-0.0673	-7471B	PPB	10 Jul 2018 08:41:05	U	SMPL
CCV33 - 1	CCV33	23146	5.0471	-7471B	PPB	10 Jul 2018 08:43:21	U	SMPL
CCB33 - 1	CCB33	166	-0.0609	-7471B	PPB	10 Jul 2018 08:45:37	U	SMPL
CRA - 1	CRA	1397	0.2128	-7471B	PPB	10 Jul 2018 08:47:53	U	SMPL
HighStd - 1	HighStd	45627	10.0441	-7471B	PPB	10 Jul 2018 08:50:08	U	SMPL
ChkStd - 1	ChkStd	31973	7.0092	-7471B	PPB	10 Jul 2018 08:52:24	U	SMPL
PB110945BL - 1	PBS	87	-0.0784	-7471B	PPB	10 Jul 2018 08:54:42	U	SMPL
PB110945BS - 1	LCSS	17168	3.7183	-7471B	PPB	10 Jul 2018 08:56:59	U	SMPL
J3874-01 - 1	2018-NYSEG-MCMaster-PH4-5(0-3)	60747	13.4050	-7471B	PPB	10 Jul 2018 08:59:16	U	SMPL
J3874-02 - 1	2018-NYSEG-MCMaster-PH4-5(3-5)	20059	4.3609	-7471B	PPB	10 Jul 2018 09:01:33	U	SMPL
J3891-01 - 1	212-72-11940-R4-VNJ-242	23868	5.2076	-7471B	PPB	10 Jul 2018 09:03:53	U	SMPL
J3891-05 - 1	72-12002-RBR-RBR-VNJ-230	21163	4.6063	-7471B	PPB	10 Jul 2018 09:06:11	U	SMPL
J3893-01 - 1	BS-S-7-SOIL	2338	0.4219	-7471B	PPB	10 Jul 2018 09:08:28	U	SMPL
CCV34 - 1	CCV34	22584	4.9222	-7471B	PPB	10 Jul 2018 09:10:45	U	SMPL
CCB34 - 1	CCB34	108	-0.0737	-7471B	PPB	10 Jul 2018 09:13:01	U	SMPL
J3896-01 - 1	JC-SOIL	8493	1.7901	-7471B	PPB	10 Jul 2018 09:15:17	U	SMPL
J3897-01 - 1	STOCK-PILE-A	19874	4.3198	-7471B	PPB	10 Jul 2018 09:17:33	U	SMPL
J3897-01DUP - 1	STOCK-PILE-ADUP	17533	3.7995	-7471B	PPB	10 Jul 2018 09:19:49	U	SMPL
J3897-01MS - 1	STOCK-PILE-AMS	35995	7.9032	-7471B	PPB	10 Jul 2018 09:22:06	U	SMPL
J3897-01MSD - 1	STOCK-PILE-AMSD	37045	8.1365	-7471B	PPB	10 Jul 2018 09:27:08	U	SMPL
J3901-03 - 1	TP-9-S	2773	0.5186	-7471B	PPB	10 Jul 2018 09:29:25	U	SMPL
J3906-01 - 1	CB-7-8-6-ST-WW-1-3	2168	0.3841	-7471B	PPB	10 Jul 2018 09:31:43	U	SMPL
J3906-02 - 1	CB-7-8-6-ST-EW-1-3	2423	0.4408	-7471B	PPB	10 Jul 2018 09:34:01	U	SMPL
J3906-03 - 1	CB-7-8-6-ST-WW-2-2.75	4880	0.9870	-7471B	PPB	10 Jul 2018 09:36:18	U	SMPL
J3906-04 - 1	CB-7-8-6-ST-EW-2-2.75	4496	0.9016	-7471B	PPB	10 Jul 2018 09:38:34	U	SMPL
CCV35 - 1	CCV35	22747	4.9584	-7471B	PPB	10 Jul 2018 09:40:50	U	SMPL
CCB35 - 1	CCB35	93	-0.0771	-7471B	PPB	10 Jul 2018 09:43:07	U	SMPL
J3906-05 - 1	CB-7-8-6-ST-BOTTOM-1-3	3576	0.6971	-7471B	PPB	10 Jul 2018 09:45:24	U	SMPL
J3906-06 - 1	CB-7-8-6-ST-WW-3-2.5	2921	0.5515	-7471B	PPB	10 Jul 2018 09:47:40	U	SMPL
J3906-07 - 1	CB-7-8-6-ST-EW-3-2.5	2708	0.5042	-7471B	PPB	10 Jul 2018 09:49:56	U	SMPL
J3906-08 - 1	STOCKPILE-NEWSAN	7352	1.5364	-7471B	PPB	10 Jul 2018 09:52:13	U	SMPL
J3906-09 - 1	STOCKPILE-FOUNDRY	7276	1.5195	-7471B	PPB	10 Jul 2018 09:54:30	U	SMPL
J3907-01 - 1	S-1	2241	0.4004	-7471B	PPB	10 Jul 2018 09:56:47	U	SMPL
J3907-02 - 1	S-2	2210	0.3935	-7471B	PPB	10 Jul 2018 09:59:04	U	SMPL
J3908-03 - 1	TP-10-S	1505	0.2368	-7471B	PPB	10 Jul 2018 10:01:22	U	SMPL
PB110948BL - 1	PBS	468	0.0063	-7471B	PPB	10 Jul 2018 10:03:40	U	SMPL
PB110948BS - 1	LCSS	15912	3.4391	-7471B	PPB	10 Jul 2018 10:05:56	U	SMPL
CCV36 - 1	CCV36	22564	4.9177	-7471B	PPB	10 Jul 2018 10:08:12	U	SMPL
CCB36 - 1	CCB36	96	-0.0764	-7471B	PPB	10 Jul 2018 10:10:28	U	SMPL
J3827-01 - 1	OR-03-070918	2381	0.4315	-7471B	PPB	10 Jul 2018 10:12:45	U	SMPL
J3827-01DUP - 1	OR-03-070918DUP	2293	0.4119	-7471B	PPB	10 Jul 2018 10:17:42	U	SMPL
J3827-01MS - 1	OR-03-070918MS	18614	4.0397	-7471B	PPB	10 Jul 2018 10:19:58	U	SMPL
J3827-01MSD - 1	OR-03-070918MSD	18962	4.1171	-7471B	PPB	10 Jul 2018 10:22:14	U	SMPL
J3903-01 - 1	KS-T-1	5726	1.1750	-7471B	PPB	10 Jul 2018 10:24:31	U	SMPL
J3903-03 - 1	KS-T-3	3433	0.6653	-7471B	PPB	10 Jul 2018 10:26:48	U	SMPL
J3914-01 - 1	B1	117644	26.0520	-7471B	PPB	10 Jul 2018 10:29:05	U	SMPL
J3914-03 - 1	B2	73794	16.3051	-7471B	PPB	10 Jul 2018 10:31:22	U	SMPL
J3914-05 - 1	B3	85690	18.9493	-7471B	PPB	10 Jul 2018 10:33:43	U	SMPL
J3914-07 - 1	NEMH	30208	6.6168	-7471B	PPB	10 Jul 2018 10:36:02	U	SMPL
CCV37 - 1	CCV37	23076	5.0315	-7471B	PPB	10 Jul 2018 10:38:21	U	SMPL
CCB37 - 1	CCB37	33	-0.0904	-7471B	PPB	10 Jul 2018 10:40:39	U	SMPL
J3914-09 - 1	SEMH	52199	11.5050	-7471B	PPB	10 Jul 2018 10:42:56	U	SMPL
J3914-11 - 1	SMH	189741	42.0776	-7471B	PPB	10 Jul 2018 10:45:12	U	SMPL
J3897-01LX5 - 1		3150	0.6024	-7471B	PPB	10 Jul 2018 10:47:31	U	SMPL
J3897-01A - 1		36703	8.0605	-7471B	PPB	10 Jul 2018 10:50:00	U	SMPL
J3827-01LX5 - 1		332	-0.0240	-7471B	PPB	10 Jul 2018 10:52:16	U	SMPL
J3827-01A - 1		19163	4.1618	-7471B	PPB	10 Jul 2018 10:54:35	U	SMPL
J3874-01DLX2 - 1		29774	6.5204	-7471B	PPB	10 Jul 2018 10:56:52	U	SMPL
CCV38 - 1	CCV38	22694	4.9466	-7471B	PPB	10 Jul 2018 10:59:10	U	SMPL
CCB38 - 1	CCB38	111	-0.0731	-7471B	PPB	10 Jul 2018 11:01:30	U	SMPL
J3914-01DLX5 - 1		24301	5.3038	-7471B	PPB	10 Jul 2018 11:03:47	U	SMPL
J3914-03DLX5 - 1		15053	3.2482	-7471B	PPB	10 Jul 2018 11:06:04	U	SMPL
J3914-05DLX5 - 1		16971	3.6745	-7471B	PPB	10 Jul 2018 11:08:22	U	SMPL
J3914-09DLX2 - 1		26700	5.8371	-7471B	PPB	10 Jul 2018 11:10:38	U	SMPL
J3914-11DLX10 - 1		19513	4.2396	-7471B	PPB	10 Jul 2018 11:12:55	U	SMPL
CCV39 - 1	CCV39	23009	5.0167	-7471B	PPB	10 Jul 2018 11:15:12	U	SMPL
CCB39 - 1	CCB39	96	-0.0764	-7471B	PPB	10 Jul 2018 11:17:29	U	SMPL