

AB96 524

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/10/2018 8:25

Test: Phenolics-my h

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.022	0.0	0.105	
ICB1	0.001	0.0	0.004	
CCV1	0.986	0.0	0.102	
CCB1	0.000	0.0	0.004	
PB110940BL	0.001	0.0	0.004	
PB110940BS	1.028	0.0	0.106	
J3762-01	0.012	0.0	0.005	
J3762-02	0.093	0.0	0.013	
J3762-03	0.011	0.0	0.005	
PB110942BL	-0.003	0.0	0.003	
PB110942BS	1.038	0.0	0.107	
J3762-07	0.009	0.0	0.004	
J3762-08	0.080	0.0	0.011	
J3762-09	0.009	0.0	0.004	
CCV2	0.987	0.0	0.102	
CCB2	0.001	0.0	0.004	

N 16
Mean 0.330
SD 0.4761
CV% 144.38

Aquakem v. 7.2AQ1

Results from time period:

Tue Jul 10 07:45:00 2018

Tue Jul 10 08:21:07 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0497	mg/l	7/10/2018 7:45:00	
0.1PPM	A	Phenolics-	P	0.1	mg/l	7/10/2018 7:45:01	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	7/10/2018 7:45:02	
0.50PPM	A	Phenolics-	P	0.4976	mg/l	7/10/2018 7:45:03	
1.0PPM	A	Phenolics-	P	1.0048	mg/l	7/10/2018 7:45:04	
2.0PPM	A	Phenolics-	P	1.9983	mg/l	7/10/2018 7:45:05	
ICV1	S	Phenolics-	P	1.0222	mg/l	7/10/2018 8:11:02	
ICB1	S	Phenolics-	P	0.0012	mg/l	7/10/2018 8:11:03	
CCV1	S	Phenolics-	P	0.9856	mg/l	7/10/2018 8:11:04	
CCB1	S	Phenolics-	P	0.0002	mg/l	7/10/2018 8:11:05	
PB110940BL	S	Phenolics-	P	0.0015	mg/l	7/10/2018 8:11:06	
PB110940BS	S	Phenolics-	P	1.0281	mg/l	7/10/2018 8:11:07	
J3762-01	S	Phenolics-	P	0.0116	mg/l	7/10/2018 8:11:09	
J3762-02	S	Phenolics-	P	0.093	mg/l	7/10/2018 8:11:10	
J3762-03	S	Phenolics-	P	0.0112	mg/l	7/10/2018 8:11:13	
PB110942BL	S	Phenolics-	P	-0.0033	mg/l	7/10/2018 8:20:57	
PB110942BS	S	Phenolics-	P	1.0379	mg/l	7/10/2018 8:20:58	
J3762-07	S	Phenolics-	P	0.0095	mg/l	7/10/2018 8:21:00	
J3762-08	S	Phenolics-	P	0.0799	mg/l	7/10/2018 8:21:02	
J3762-09	S	Phenolics-	P	0.0094	mg/l	7/10/2018 8:21:03	
CCV2	S	Phenolics-	P	0.9868	mg/l	7/10/2018 8:21:05	
CCB2	S	Phenolics-	P	0.0011	mg/l	7/10/2018 8:21:06	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

7/10/2018 7:45


Test Phenolics-

Accepted 7/10/2018 7:45

Factor 10.06
 Bias 0.004

Coeff. of det. 0.999989

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0497	0.0500	
2	Phenol-2	0.013	0.1000	0.1000	
3	Phenol-2	0.028	0.2497	0.2500	
4	Phenol-2	0.053	0.4976	0.5000	
5	Phenol-2	0.103	1.0048	1.0000	
6	Phenol-2	0.202	1.9983	2.0000	