

245.1


LB96537

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35047



A= 0.0000e+000

B= 2.8134e-004

C= 1.1604e-002

Rho= 0.9997057

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-111	0.000	-111				
0.05	0.050	0.054	0.004	150	0.0 %	150				
0.2	0.200	0.217	0.017	731	0.0 %	731				
2.5	2.500	2.458	-0.042	8695	0.0 %	8695				
5.0	5.000	4.976	-0.024	17645	0.0 %	17645				
7.5	7.500	7.693	0.193	27303	0.0 %	27303				
10.0	10.000	9.872	-0.128	35047	0.0 %	35047				

LB96537 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Jul 2018 15:43:03

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-111	-	0.0000245.1	PPB	10 Jul 2018 15:46:37	S	Std
0.05 - 1	50.05	150	-	0.0500245.1	PPB	10 Jul 2018 15:48:46	S	Std
0.2 - 1	50.2	731	-	0.2000245.1	PPB	10 Jul 2018 15:50:54	S	Std
2.5 - 1	52.5	8695	-	2.5000245.1	PPB	10 Jul 2018 15:53:03	S	Std
5.0 - 1	55	17645	-	5.0000245.1	PPB	10 Jul 2018 15:55:11	S	Std
7.5 - 1	57.5	27303	-	7.5000245.1	PPB	10 Jul 2018 15:57:18	S	Std
10.0 - 1	510	35047	-	10.0000245.1	PPB	10 Jul 2018 15:59:26	S	Std
ICV42 - 1	ICV42	14536	4.1012	-245.1	PPB	10 Jul 2018 16:06:48	U	SMPL
ICB42 - 1	ICB42	-245	-0.0573	-245.1	PPB	10 Jul 2018 16:08:54	U	SMPL
CCV44 - 1	CCV44	17659	4.9798	-245.1	PPB	10 Jul 2018 16:11:01	U	SMPL
CCB44 - 1	CCB44	-281	-0.0675	-245.1	PPB	10 Jul 2018 16:13:07	U	SMPL
CRA - 1	CRA	124	0.0465	-245.1	PPB	10 Jul 2018 16:15:15	U	SMPL
HighStd - 1	HighStd	35477	9.9927	-245.1	PPB	10 Jul 2018 16:17:21	U	SMPL
ChkStd - 1	ChkStd	25309	7.1321	-245.1	PPB	10 Jul 2018 16:19:28	U	SMPL
PB110934BL - 1	PBW	-353	-0.0877	-245.1	PPB	10 Jul 2018 16:21:37	U	SMPL
PB110934BS - 1	LCSW	12925	3.6479	-245.1	PPB	10 Jul 2018 16:23:46	U	SMPL
J3876-01 - 1	070318-DBRI-E-D-M	-353	-0.0877	-245.1	PPB	10 Jul 2018 16:25:54	U	SMPL
J3876-02 - 1	070318-DBRI-E-U-M	-237	-0.0551	-245.1	PPB	10 Jul 2018 16:28:03	U	SMPL
J3876-03 - 1	070318-TIPC-E-D-M	-222	-0.0509	-245.1	PPB	10 Jul 2018 16:30:11	U	SMPL
J3876-04 - 1	070318-TIPC-E-U-M	-227	-0.0523	-245.1	PPB	10 Jul 2018 16:32:19	U	SMPL
J3890-01 - 1	070518-DBRI-E-D-M	-240	-0.0559	-245.1	PPB	10 Jul 2018 16:34:26	U	SMPL
CCV45 - 1	CCV45	17934	5.0572	-245.1	PPB	10 Jul 2018 16:36:33	U	SMPL
CCB45 - 1	CCB45	-274	-0.0655	-245.1	PPB	10 Jul 2018 16:38:40	U	SMPL
J3890-02 - 1	070518-DBRI-E-U-M	-217	-0.0494	-245.1	PPB	10 Jul 2018 16:40:47	U	SMPL
J3890-03 - 1	070518-TIPC-E-D-M	-170	-0.0362	-245.1	PPB	10 Jul 2018 16:42:53	U	SMPL
J3890-04 - 1	070518-TIPC-E-U-M	-198	-0.0441	-245.1	PPB	10 Jul 2018 16:45:00	U	SMPL
J3890-04DUP - 1	070518-TIPC-E-U-MDUP	-211	-0.0478	-245.1	PPB	10 Jul 2018 16:47:07	U	SMPL
J3890-04MS - 1	070518-TIPC-E-U-MMS	15864	4.4748	-245.1	PPB	10 Jul 2018 16:49:15	U	SMPL
J3890-04MSD - 1	070518-TIPC-E-U-MMSD	15284	4.3116	-245.1	PPB	10 Jul 2018 16:51:23	U	SMPL
J3890-04LX5 - 1		-263	-0.0624	-245.1	PPB	10 Jul 2018 16:53:30	U	SMPL
J3890-04A - 1		14399	4.0626	-245.1	PPB	10 Jul 2018 16:55:39	U	SMPL
CCV46 - 1	CCV46	17918	5.0527	-245.1	PPB	10 Jul 2018 16:57:48	U	SMPL
CCB46 - 1	CCB46	-285	-0.0686	-245.1	PPB	10 Jul 2018 16:59:55	U	SMPL