



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96542

Slope : 98.9

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/10/2018	15:15
2	CAL2	1	Water			21.6	7.00	07/10/2018	15:17
3	CAL3	1	Water			21.7	10.03	07/10/2018	15:19
4	ICV	1	Water			21.5	7.00	07/10/2018	15:21
5	CCV1	1	Water			21.8	2.00	07/10/2018	15:23
6	J3928-04	1	Water			18.8	7.27	07/10/2018	15:25
7	J3928-04DUP	1	Water			18.9	7.28	07/10/2018	15:27
8	CCV2	1	Water			21.6	12.00	07/10/2018	15:29