



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96543

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/10/2018	15:15
2	CAL2	1	Water			21.6	7.00	07/10/2018	15:17
3	CAL3	1	Water			21.7	10.03	07/10/2018	15:19
4	ICV	1	Water			21.5	7.00	07/10/2018	15:21
5	CCV1	1	Water			21.7	2.00	07/10/2018	16:45
6	J3824-02	1	Solid	20.02	20	24.5	8.03	07/10/2018	16:47
7	J3824-02DUP	1	Solid	20.02	20	24.6	8.04	07/10/2018	16:49
8	J3828-02	1	Solid	20.03	20	24.3	8.63	07/10/2018	16:51
9	J3828-04	1	Solid	20.02	20	24.0	7.67	07/10/2018	16:53
10	J3828-06	1	Solid	20.03	20	23.3	7.82	07/10/2018	16:55
11	J3832-02	1	Solid	20.04	20	24.6	7.52	07/10/2018	16:57
12	J3918-02	1	Solid	20.03	20	22.7	8.85	07/10/2018	16:59
13	J3924-02	1	Solid	20.02	20	22.5	8.78	07/10/2018	17:01
14	J3926-02	1	Solid	20.04	20	22.7	9.37	07/10/2018	17:03
15	J3927-04	1	Solid	20.02	20	24.3	7.80	07/10/2018	17:05
16	CCV2	1	Water			21.7	12.00	07/10/2018	17:07
17	J3928-02	1	Solid	20.02	20	22.7	8.96	07/10/2018	17:09
18	J3930-04	1	Solid	20.03	20	24.5	6.35	07/10/2018	17:11
19	CCV3	1	Water			21.8	2.00	07/10/2018	17:13

WORKLIST(Hardcopy Internal Chain)

LB96543

WorkList Name : corrosivity071018

WorkList ID : 114285

Date : 7/10/2018 2:50:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2018	Solid	J3824-02	Corrosivity	Cool 4 deg C	PSEG05	K33	HD-02-071018	07/10/2018	9045C
07/10/2018	Solid	J3828-02	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-071018-A	07/10/2018	9045C
07/10/2018	Solid	J3828-04	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-071018-B	07/10/2018	9045C
07/10/2018	Solid	J3828-06	Corrosivity	Cool 4 deg C	PSEG05	K31	EO-01-071018-C	07/10/2018	9045C
07/10/2018	Solid	J3832-02	Corrosivity	Cool 4 deg C	PSEG05	K31	TR-05-071018	07/10/2018	9045C
07/10/2018	Solid	J3918-02	Corrosivity	Cool 4 deg C	PSEG01	K22	VNJ-204-255-ETS	07/10/2018	9045C
07/10/2018	Solid	J3924-02	Corrosivity	Cool 4 deg C	PSEG01	K23	RO-1-4	07/10/2018	9045C
07/10/2018	Solid	J3926-02	Corrosivity	Cool 4 deg C	PSEG01	K31	TRAP-BAG	07/10/2018	9045C
07/10/2018	Solid	J3927-04	Corrosivity	Cool 4 deg C	PSEG01	K32	CS-RBR-200031	07/10/2018	9045C
07/09/2018	Solid	J3928-02	Corrosivity	Cool 4 deg C	SEAC01	K12	SOIL-CUTTINGS	07/09/2018	9045C
07/10/2018	Solid	J3930-04	Corrosivity	Cool 4 deg C	PSEG01	K33	NYS-RBR-200030	07/10/2018	9045C

Date/Time 07/10/18 4:30 PM

Received by: SP

Relinquished by: CP

Date/Time 07/10/18 5:30 PM

Received by: CP

Relinquished by: SP