



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96544

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	07/10/2018	10:10
2	J3911-02		1	5	>100.0	>212.0	>212.0	07/10/2018	10:32
3	J3911-02DUP		1	5	>100.0	>212.0	>212.0	07/10/2018	10:55
4	J3911-04		1	7	>100.0	>212.0	>212.0	07/10/2018	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96544

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71018-1

WorkList ID : 114274

Date : 7/10/2018 9:12:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2018	Water	J3911-02	Flash Point	Cool 4 deg C	PSEG01	K12	RB47196RT	07/09/2018	1010A
07/12/2018	Water	J3911-04	Flash Point	Cool 4 deg C	PSEG01	K12	VB21029	07/09/2018	1010A

7/10/18 9:30 AM

Date/Time

Received by:

Relinquished by:

KS
CP

7/10/18 11:25 AM

Date/Time

Received by:

Relinquished by:

CP
KS