

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/11/2018

TEMP1 IN: 103 07/11/2018 07:00 TEMP OUT: 104 07/11/2018 08:05

TEMP2 IN: 104 07/11/2018 08:45 TEMP OUT: 104 07/11/2018 10:00

TEMP3 IN: 105 07/11/2018 11:15 TEMP OUT: 105 07/11/2018 12:50

TEMP4 IN: 105 07/11/2018 13:30 TEMP OUT: 105 07/11/2018 15:05

Run Number: LB96550

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96550BL	LB96550BL	1.4030	1.4030	100	1.4030	1.4030	1.4030	0.0000	0
2	LB96550BS	LB96550BS	1.4159	1.4159	100	1.4705	1.4705	1.4705	0.0546	546
3	J3921-02	KY038PS027-180709	1.4125	1.4125	700	1.4140	1.4140	1.4140	0.0015	2.1
4	J3931-01	070918-DBRI-E-D-B	1.3991	1.3991	250	1.4140	1.4140	1.4140	0.0149	59.6
5	J3931-01DUP	070918-DBRI-E-D-BDUP	1.4025	1.4025	250	1.4170	1.4170	1.4170	0.0145	58
6	J3931-02	070918-DBRI-E-D-M	1.4111	1.4111	400	1.4234	1.4234	1.4234	0.0123	30.8
7	J3931-03	070918-DBRI-E-D-S	1.4089	1.4089	400	1.4194	1.4194	1.4194	0.0105	26.3
8	J3931-04	070918-DBRI-E-U-B	1.4071	1.4071	200	1.4238	1.4238	1.4238	0.0167	83.5
9	J3931-05	070918-DBRI-E-U-M	1.4048	1.4048	400	1.4153	1.4153	1.4153	0.0105	26.3
10	J3931-06	070918-DBRI-E-U-S	1.4031	1.4031	400	1.4113	1.4113	1.4113	0.0082	20.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS071118

WorkList ID : 114299

Date : 7/11/2018 7:53:32 AM

2196550

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2018	Water	J3921-02 C	TSS	Cool 4 deg C	SHAW22	K23	KY038PS027-180709	07/09/2018	SM2540 D
07/16/2018	Water	J3931-01 C	TSS	Cool 4 deg C	HDR103	K31	070918-DBRI-E-D-B	07/09/2018	SM2540 D
07/16/2018	Water	J3931-02 C	TSS	Cool 4 deg C	HDR103	K31	070918-DBRI-E-D-M	07/09/2018	SM2540 D
07/16/2018	Water	J3931-03 C	TSS	Cool 4 deg C	HDR103	K31	070918-DBRI-E-D-S	07/09/2018	SM2540 D
07/16/2018	Water	J3931-04 C	TSS	Cool 4 deg C	HDR103	K31	070918-DBRI-E-U-B	07/09/2018	SM2540 D
07/16/2018	Water	J3931-05 C	TSS	Cool 4 deg C	HDR103	K31	070918-DBRI-E-U-M	07/09/2018	SM2540 D
07/16/2018	Water	J3931-06 C	TSS	Cool 4 deg C	HDR103	K31	070918-DBRI-E-U-S	07/09/2018	SM2540 D

FP-ID-9823888

Date/Time

Received by:

Relinquished by:

07/11/18 08:00
AS
M-S

Date/Time

Received by:

Relinquished by:

07/11/18 12:00
M.S.P.
AS