



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96557

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	5	27.0	80.6	80.9	07/11/2018	09:40
2	J3904-02		1	6	>100.0	>212.0	>212.0	07/11/2018	10:05
3	J3904-02DUP		1	6	>100.0	>212.0	>212.0	07/11/2018	10:35
4	J3904-04		1	5	>100.0	>212.0	>212.0	07/11/2018	10:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96557

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71018

WorkList ID : 114273

Date : 7/10/2018 9:10:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2018	Solid	J3904-02	Flash Point	Cool 4 deg C	PSEG01	K11	RAR-DRUM-2	07/09/2018	1010A
07/12/2018	Solid	J3904-04	Flash Point	Cool 4 deg C	PSEG01	K11	REFURBISHED-DRUM-2-3	07/09/2018	1010A

Date/Time 7/11/18 9:15 AM
Received by: KS
Relinquished by: CP

Date/Time 7/11/18 11:00 AM
Received by: CP
Relinquished by: KS