



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96559

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	5	27.0	80.6	80.9	07/11/2018	11:25
2	J3928-04		1	7	>100.0	>212.0	>212.0	07/11/2018	11:50
3	J3928-04DUP		1	7	>100.0	>212.0	>212.0	07/11/2018	12:15
4	J3934-02		1	7	>100.0	>212.0	>212.0	07/11/2018	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96559

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71118

WorkList ID : 114323

Date : 7/11/2018 9:46:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/22/2018	Water	J3928-04 C	Flash Point	Cool 4 deg C	SEAC01	K12	GROUNDWATER	07/09/2018	1010A
07/15/2018	Water	J3934-02 C	Flash Point	Cool 4 deg C	PSEG01	K33	TS-BASELINE-WATER	07/11/2018	1010A

Date/Time 7/11/18 11:00 AM
Received by: KS
Relinquished by: CP

Date/Time 7/11/18 12:45 PM
Received by: CP
Relinquished by: KS