



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96562

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/11/2018	11:25
2	CAL2	1	Water			21.6	7.00	07/11/2018	11:27
3	CAL3	1	Water			21.7	10.03	07/11/2018	11:29
4	ICV	1	Water			21.8	7.00	07/11/2018	11:31
5	CCV1	1	Water			21.7	2.00	07/11/2018	11:33
6	J3934-01	1	Water			22.0	7.55	07/11/2018	11:35
7	J3934-01DUP	1	Water			22.1	7.56	07/11/2018	11:37
8	CCV2	1	Water			21.5	12.00	07/11/2018	11:39

WORKLIST(Hardcopy Internal Chain)

LB 96562

WorkList Name : pH071118

WorkList ID : 114329

Date : 7/11/2018 11:07:37 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2018	Water	J3934-01	pH	Cool 4 deg C	PSEG01	K33	TS-BASELINE-WATER	07/11/2018	9040C

Date/Time 07/11/18 11:10
 Received by: SP
 Relinquished by: CP

Date/Time 07/11/18 11:25
 Received by: CP
 Relinquished by: SP