



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB96563

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP66045
Residual chlorine std, 0PPM	WP66035
Residual chlorine std, 0.1PPM	WP66037
Residual chlorine LOQ, 0.2PPM	WP66053
Residual chlorine std, 0.2PPM	WP66041
Residual chlorine std, and CCV , 0.4PPM	WP66043
Residual chlorine std, 1.6PPM	WP66048
Residual chlorine ICV-LCS, 0.4PPM	WP66051
Residual chlorine LOD, 0.05PPM	WP66052
Total Chlorine Powder Pillows	W1960

Intercept: 0.0000

Slope: 1.0000

Regression: 1.000000

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	07/11/2018	11:30
2	CAL2	0.1	1	0.000	0.100	0.100	07/11/2018	11:30
3	CAL3	0.2	1	0.000	0.200	0.200	07/11/2018	11:31
4	CAL4	0.4	1	0.000	0.400	0.400	07/11/2018	11:32
5	CAL5	0.8	1	0.000	0.800	0.800	07/11/2018	11:33
6	CAL6	1.6	1	0.000	1.600	1.600	07/11/2018	11:34



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB96563

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.210	0.210	0.210	07/11/2018	11:35
2	ICB				1	0.000	0.000	0.000	0.000	07/11/2018	11:36
3	CCV1			0.4	1	0.000	0.430	0.430	0.430	07/11/2018	11:37
4	CCB1				1	0.000	0.000	0.000	0.000	07/11/2018	11:38
5	LB96563BL	50	50		1	0.000	0.000	0.000	0.000	07/11/2018	11:39
6	LB96563BS	50	50	0.4	1	0.000	0.420	0.420	0.420	07/11/2018	11:40
7	J3762-07	50	50		1	0.000	0.070	0.070	0.070	07/11/2018	11:41
8	J3762-08	50	50		1	0.000	0.240	0.240	0.240	07/11/2018	11:42
9	J3762-09	50	50		1	0.000	0.060	0.060	0.060	07/11/2018	11:43
10	CCV2			0.4	1	0.000	0.400	0.400	0.400	07/11/2018	11:44
11	CCB2				1	0.000	0.000	0.000	0.000	07/11/2018	11:45

LB96563

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RCL071118

WorkList ID : 114298

Date : 7/11/2018 7:52:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/28/2018	Water	J3762-07	Residual Chlorine	Cool 4 deg C	CHEM02	QA Office	LOD-MDL-WATER-01-QT3-20	06/28/2018	SM4500 Cl G
06/28/2018	Water	J3762-08	Residual Chlorine	Cool 4 deg C	CHEM02	QA Office	LOQ-WATER-02-QT3-2018	06/28/2018	SM4500 Cl G
06/28/2018	Water	J3762-09	Residual Chlorine	Cool 4 deg C	CHEM02	QA Office	MDL-WATER-03-QT3-2018	06/28/2018	SM4500 Cl G

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: