



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96573

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	07/11/2018	16:50
2	CAL2	1	Water			21.7	7.00	07/11/2018	16:52
3	CAL3	1	Water			21.8	10.03	07/11/2018	16:54
4	ICV	1	Water			21.8	7.00	07/11/2018	16:56
5	CCV1	1	Water			21.6	2.00	07/11/2018	16:58
6	J3825-02	1	Solid	20.02	20	25.1	7.88	07/11/2018	17:00
7	J3825-02DUP	1	Solid	20.04	20	25.2	7.89	07/11/2018	17:02
8	J3825-04	1	Solid	20.02	20	24.8	8.62	07/11/2018	17:04
9	J3825-06	1	Solid	20.03	20	24.2	8.96	07/11/2018	17:06
10	J3825-08	1	Solid	20.02	20	24.6	8.65	07/11/2018	17:08
11	J3825-10	1	Solid	20.04	20	25.0	8.44	07/11/2018	17:10
12	J3863-01	1	Solid	20.02	20	23.2	12.92	07/11/2018	17:12
13	J3863-02	1	Solid	20.02	20	22.9	10.33	07/11/2018	17:14
14	J3863-03	1	Solid	20.02	20	23.2	9.80	07/11/2018	17:16
15	J3947-02	1	Solid	20.04	20	24.1	7.91	07/11/2018	17:18
16	CCV2	1	Water			21.8	12.00	07/11/2018	17:20
17	J3947-04	1	Solid	20.02	20	23.9	7.13	07/11/2018	17:22
18	J3947-06	1	Solid	20.04	20	24.8	7.77	07/11/2018	17:24
19	J3947-06DUP	1	Solid	20.04	20	24.9	7.78	07/11/2018	17:26
20	J3947-08	1	Solid	20.02	20	24.9	7.22	07/11/2018	17:28

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			21.7	2.00	07/11/2018	17:30

WORKLIST(Hardcopy Internal Chain)

LB 96573

WorkList Name : corrosivity071118

WorkList ID : 114340

Date : 7/11/2018 4:29:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2018	Solid	J3825-02	Corrosivity	Cool 4 deg C	PSEG05	K33	CL-02-071118-A	07/11/2018	9045C
07/11/2018	Solid	J3825-04	Corrosivity	Cool 4 deg C	PSEG05	K33	CL-02-071118-B	07/11/2018	9045C
07/11/2018	Solid	J3825-06	Corrosivity	Cool 4 deg C	PSEG05	K33	CL-02-071118-C	07/11/2018	9045C
07/11/2018	Solid	J3825-08	Corrosivity	Cool 4 deg C	PSEG05	K33	CL-02-071118-D	07/11/2018	9045C
07/11/2018	Solid	J3825-10	Corrosivity	Cool 4 deg C	PSEG05	K33	CL-02-071118-E	07/11/2018	9045C
06/12/2018	Solid	J3863-01	Corrosivity	Cool 4 deg C	ATCE02	K23	180612NHL14S-01-LCC-PAIN	06/12/2018	9045C
06/12/2018	Solid	J3863-02	Corrosivity	Cool 4 deg C	ATCE02	K23	180612NHL14S-02-LCC-POL	06/12/2018	9045C
06/12/2018	Solid	J3863-03	Corrosivity	Cool 4 deg C	ATCE02	K23	180612NHL14S-03-LCC-PPE	06/12/2018	9045C
07/11/2018	Solid	J3947-02	Corrosivity	Cool 4 deg C	PSEG01	K41	PAR-1	07/11/2018	9045C
07/11/2018	Solid	J3947-04	Corrosivity	Cool 4 deg C	PSEG01	K41	PAR-2	07/11/2018	9045C
07/11/2018	Solid	J3947-06	Corrosivity	Cool 4 deg C	PSEG01	K41	PAR-3	07/11/2018	9045C
07/11/2018	Solid	J3947-08	Corrosivity	Cool 4 deg C	PSEG01	K41	PAR-4	07/11/2018	9045C

Date/Time 07/11/18 4:35

Received by: SP

Relinquished by: CP

Date/Time 07/11/18 5:40

Received by: CP

Relinquished by: SP