



PERCENT SOLID

Analyst Name: jignesh
Date: 7/12/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 07/11/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:20
Out Date: 07/12/2018
Weight Check 1.0g: 1.00 g
Weight Check 10g: 10.00 g
BalanceID: M Sc-1

QC: LB96574

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J3633-01	A-1 (0-6)	1	1.14	9.96	8.48	83.2
J3633-02	A-1 (5.5)	2	1.13	9.66	8.66	88.3
J3633-03	B-1 (0-6)	3	1.16	9.97	8.44	82.6
J3633-04	B-1 (3.0)	4	1.19	9.61	8.69	89.1
J3633-05	C-1 (0-3)	5	1.11	9.65	8.39	85.2
J3633-06	C-1 (2.0)	6	1.15	9.82	8.63	86.3
J3825-01	CL-02-071118-A	7	1.14	9.59	9.22	95.6
J3825-03	CL-02-071118-B	8	1.19	9.66	9.20	94.6
J3825-05	CL-02-071118-C	9	1.17	9.59	9.32	96.8
J3825-07	CL-02-071118-D	10	1.13	9.89	9.45	95
J3825-09	CL-02-071118-E	11	1.16	9.97	9.41	93.6
J3943-01	C-1	12	1.14	9.56	9.48	99
J3944-01	FOUNDRY-WEST	13	1.13	9.81	9.51	96.5
J3944-02	FOUNDRY-COMP	14	1.16	9.69	8.27	83.4
J3944-03	FOUNDRY-EAST	15	1.17	9.67	9.32	95.9
J3947-01	PAR-1	16	1.16	9.95	9.21	91.6
J3947-02	PAR-1	17	1.16	9.95	9.21	91.6
J3947-03	PAR-2	18	1.13	9.70	6.50	62.7
J3947-04	PAR-2	19	1.13	9.70	6.50	62.7
J3947-05	PAR-3	20	1.19	9.88	5.67	51.6
J3947-06	PAR-3	21	1.19	9.88	5.67	51.6
J3947-07	PAR-4	22	1.17	9.96	7.41	71
J3947-08	PAR-4	23	1.17	9.96	7.41	71
J3947-09	OLD-FOOTING-1	24	1.00	2.00	2.00	100
J3947-11	OLD-FOOTING-2	25	1.00	2.00	2.00	100
J3947-13	OLD-FOOTING-3	26	1.00	2.00	2.00	100
J3947-15	CABLE-PATH	27	1.00	2.00	2.00	100
J3947-17	NEW-FOOTING-1	28	1.00	2.00	2.00	100
J3947-19	NEW-FOOTING-2	29	1.00	2.00	2.00	100
J3947-21	NEW-FOOTING-3	30	1.00	2.00	2.00	100
J3948-01	2018-NYSEG-CLARK-ST-AREA	31	1.10	9.61	8.46	86.5
J3948-03	2018-NYSEG-CLARK-ST-AREA	32	1.12	9.94	7.93	77.2
J3948-05	2018-NYSEG-CLARK-ST-AREA	33	1.14	9.85	8.58	85.4

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

1896574

WorkList Name : %1-071118

WorkList ID : 114293

Date : 7/11/2018 7:30:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/18/2018	Solid	J3633-01	Percent Solids	Cool 4 deg C	EART13	K42	A-1(0-6)	07/09/2018	Chemtech -SO
07/18/2018	Solid	J3633-02	Percent Solids	Cool 4 deg C	EART13	K42	A-1(5.5)	07/09/2018	Chemtech -SO
07/18/2018	Solid	J3633-03	Percent Solids	Cool 4 deg C	EART13	K42	B-1(0-6)	07/09/2018	Chemtech -SO
07/18/2018	Solid	J3633-04	Percent Solids	Cool 4 deg C	EART13	K42	B-1(3.0)	07/09/2018	Chemtech -SO
07/18/2018	Solid	J3633-05	Percent Solids	Cool 4 deg C	EART13	K42	C-1(0-3)	07/09/2018	Chemtech -SO
07/18/2018	Solid	J3633-06	Percent Solids	Cool 4 deg C	EART13	K42	C-1(2.0)	07/09/2018	Chemtech -SO
07/15/2018	Solid	J3825-01	Percent Solids	Cool 4 deg C	PSEG05	K33	CL-02-071118-A	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3825-03	Percent Solids	Cool 4 deg C	PSEG05	K33	CL-02-071118-B	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3825-05	Percent Solids	Cool 4 deg C	PSEG05	K33	CL-02-071118-C	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3825-07	Percent Solids	Cool 4 deg C	PSEG05	K33	CL-02-071118-D	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3825-09	Percent Solids	Cool 4 deg C	PSEG05	K33	CL-02-071118-E	07/11/2018	Chemtech -SO
07/20/2018	Solid	J3943-01	Percent Solids	Cool 4 deg C	EARTH03	K51	C-1	07/10/2018	Chemtech -SO
07/15/2018	Solid	J3944-01	Percent Solids	Cool 4 deg C	LIRO01	K51	FOUNDRY-WEST	07/09/2018	Chemtech -SO
07/15/2018	Solid	J3944-02	Percent Solids	Cool 4 deg C	LIRO01	K51	FOUNDRY-COMP	07/09/2018	Chemtech -SO
07/15/2018	Solid	J3944-03	Percent Solids	Cool 4 deg C	LIRO01	K51	FOUNDRY-EAST	07/09/2018	Chemtech -SO
07/15/2018	Solid	J3947-01	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-1	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-02	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-1	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-03	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-2	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-04	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-2	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-05	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-3	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-06	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-3	07/11/2018	Chemtech -SO

Date/Time 07-11-18 3:50 PM

Received by: JA

Relinquished by: CB

Date/Time 07-11-18

Received by:

Relinquished by: JQ

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-071118

WorkList ID : 114293

Date : 7/11/2018 7:30:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/15/2018	Solid	J3947-07	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-4	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-08	Percent Solids	Cool 4 deg C	PSEG01	K41	PAR-4	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-09	Percent Solids	Cool 4 deg C	PSEG01	K41	OLD-FOOTING-1	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-11	Percent Solids	Cool 4 deg C	PSEG01	K41	OLD-FOOTING-2	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-13	Percent Solids	Cool 4 deg C	PSEG01	K41	OLD-FOOTING-3	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-15	Percent Solids	Cool 4 deg C	PSEG01	K41	CABLE-PATH	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-17	Percent Solids	Cool 4 deg C	PSEG01	K41	NEW-FOOTING-1	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-19	Percent Solids	Cool 4 deg C	PSEG01	K41	NEW-FOOTING-2	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3947-21	Percent Solids	Cool 4 deg C	PSEG01	K41	NEW-FOOTING-3	07/11/2018	Chemtech -SO
07/15/2018	Solid	J3948-01	Percent Solids	Cool 4 deg C	PARS02	K61	2018-NYSEG-CLARK-ST-ARE	07/10/2018	Chemtech -SO
07/15/2018	Solid	J3948-03	Percent Solids	Cool 4 deg C	PARS02	K61	2018-NYSEG-CLARK-ST-ARE	07/10/2018	Chemtech -SO
07/15/2018	Solid	J3948-05	Percent Solids	Cool 4 deg C	PARS02	K61	2018-NYSEG-CLARK-ST-ARE	07/10/2018	Chemtech -SO

Date/Time 07-11-18 3:30 PM
 Received by: JO
 Relinquished by: JO

Date/Time 07-11-18 5:30 PM
 Received by: JO
 Relinquished by: JO