



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96584

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	07/11/2018	16:50
2	CAL2	1	Water			21.7	7.00	07/11/2018	16:52
3	CAL3	1	Water			21.8	10.03	07/11/2018	16:54
4	ICV	1	Water			21.8	7.00	07/11/2018	16:56
5	CCV1	1	Water			21.6	2.00	07/11/2018	17:50
6	J3863-04	1	Water			22.6	12.02	07/11/2018	17:52
7	J3863-05	1	Water			22.5	5.86	07/11/2018	17:54
8	J3947-28	1	Water			24.9	6.09	07/11/2018	17:56
9	J3947-28DUP	1	Water			24.8	6.10	07/11/2018	17:58
10	CCV2	1	Water			21.7	12.00	07/11/2018	18:00

WORKLIST(Hardcopy Internal Chain)

LB 96584

WorkList Name : corrosivity071118-2

WorkList ID : 114342

Date : 7/11/2018 4:29:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/05/2018	Water	J3863-04	Corrosivity	Cool 4 deg C	ATCE02	K23	180612NHL14S-04-LCC-WAS	07/05/2018	9040C
07/05/2018	Water	J3863-05	Corrosivity	Cool 4 deg C	ATCE02	K23	180612NHL14S-05-LCC-HAN	07/05/2018	9040C
07/11/2018	Water	J3947-28	Corrosivity	Cool 4 deg C	PSEG01	K41	PAR-1-OILY-WATER	07/11/2018	9040C

Date/Time 07/11/18 5:45
 Received by: SP
 Relinquished by: CP

Date/Time 07/11/18 6:10
 Received by: CP
 Relinquished by: SP