



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96585

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	07/11/2018	16:50
2	CAL2	1	Water			21.7	7.00	07/11/2018	16:52
3	CAL3	1	Water			21.8	10.03	07/11/2018	16:54
4	ICV	1	Water			21.8	7.00	07/11/2018	16:56
5	CCV1	1	Water			21.6	2.00	07/11/2018	17:50
6	J3947-27	1	Water			24.9	6.09	07/11/2018	17:56
7	J3947-27DUP	1	Water			24.8	6.10	07/11/2018	17:58
8	CCV2	1	Water			21.7	12.00	07/11/2018	18:00

WORKLIST(Hardcopy Internal Chain)

LB96 SSS

WorkList Name : pH071118-1

WorkList ID : 114341

Date : 7/11/2018 4:29:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/11/2018	Water	J3947-27	pH	Cool 4 deg C	PSEG01	K41	PAR-1-OILY-WATER	07/11/2018	9040C

Date/Time 07/11/18 5:45
 Received by: SP
 Relinquished by: CP

Date/Time 07/11/18 5:45
 Received by: CP
 Relinquished by: SP