



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96586

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	07/11/2018	16:50
2	CAL2	1	Water			21.7	7.00	07/11/2018	16:52
3	CAL3	1	Water			21.8	10.03	07/11/2018	16:54
4	ICV	1	Water			21.8	7.00	07/11/2018	16:56
5	CCV1	1	Water			21.5	2.00	07/11/2018	18:05
6	J3941-01	1	Water			18.4	4.97	07/11/2018	18:07
7	J3941-01DUP	1	Water			18.5	4.98	07/11/2018	18:09
8	CCV2	1	Water			21.7	12.00	07/11/2018	18:11

LB 96586

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pHJ3941

WorkList ID : 114372

Date : 7/11/2018 5:05:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/10/2018	Water	J3941-01 E	pH	Cool 4 deg C	PARS02	K43	SES-MS-PH4-COA-7-18	07/10/2018	SM 4500-PH B

Date/Time 07/11/18 5:20

Received by: SP

Relinquished by: CP

Date/Time 07/11/18 6:20

Received by: CP

Relinquished by: SP