


7470A

LB 96587
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39011



A= 0.0000e+000

B= 2.5692e-004

C= 1.2611e-002

Rho= 0.9999679

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-33	0.000	-33				
0.2	0.200	0.233	0.033	856	0.0 %	856				
2.5	2.500	2.491	-0.009	9647	0.0 %	9647				
5.0	5.000	4.948	-0.052	19209	0.0 %	19209				
7.5	7.500	7.489	-0.011	29102	0.0 %	29102				
10.0	10.000	10.035	0.035	39011	0.0 %	39011				

LB96587 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jul 2018 09:48:00

Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-33	-	0.00007470A	PPB	12 Jul 2018 09:50:47	S	Std
0.2 - 1	50.2	856	-	0.20007470A	PPB	12 Jul 2018 09:52:55	S	Std
2.5 - 1	52.5	9647	-	2.50007470A	PPB	12 Jul 2018 09:55:03	S	Std
5.0 - 1	55	19209	-	5.00007470A	PPB	12 Jul 2018 09:57:11	S	Std
7.5 - 1	57.5	29102	-	7.50007470A	PPB	12 Jul 2018 09:59:19	S	Std
10.0 - 1	510	39011	-	10.00007470A	PPB	12 Jul 2018 10:03:53	S	Std
ICV45 - 1	ICV45	13979	3.6040	-7470A	PPB	12 Jul 2018 10:06:25	U	SMPL
ICB45 - 1	ICB45	-160	-0.0285	-7470A	PPB	12 Jul 2018 10:08:32	U	SMPL
CCV57 - 1	CCV57	18991	4.8917	-7470A	PPB	12 Jul 2018 10:10:38	U	SMPL
CCB57 - 1	CCB57	-172	-0.0316	-7470A	PPB	12 Jul 2018 10:12:45	U	SMPL
CRA - 1	CRA	852	0.2315	-7470A	PPB	12 Jul 2018 10:14:52	U	SMPL
HighStd - 1	HighStd	39047	10.0444	-7470A	PPB	12 Jul 2018 10:16:58	U	SMPL
ChkStd - 1	ChkStd	26951	6.9368	-7470A	PPB	12 Jul 2018 10:19:05	U	SMPL
PB111013BL - 1	PBW	-228	-0.0460	-7470A	PPB	12 Jul 2018 10:21:13	U	SMPL
PB111013BS - 1	LCSW	13753	3.5460	-7470A	PPB	12 Jul 2018 10:23:22	U	SMPL
PB111013TB - 1	PB110982TB	-136	-0.0223	-7470A	PPB	12 Jul 2018 10:25:30	U	SMPL
J3677-13 - 1	WASTE-CHARACTERIZATION-COMP	-14	0.0090	-7470A	PPB	12 Jul 2018 10:27:37	U	SMPL
J3824-02 - 1	HD-02-071018	-43	0.0016	-7470A	PPB	12 Jul 2018 10:29:46	U	SMPL
J3828-02 - 1	EO-01-071018-A	-32	0.0044	-7470A	PPB	12 Jul 2018 10:31:54	U	SMPL
J3828-04 - 1	EO-01-071018-B	7	0.0144	-7470A	PPB	12 Jul 2018 10:34:01	U	SMPL
CCV58 - 1	CCV58	19067	4.9112	-7470A	PPB	12 Jul 2018 10:36:08	U	SMPL
CCB58 - 1	CCB58	-149	-0.0257	-7470A	PPB	12 Jul 2018 10:38:14	U	SMPL
J3828-06 - 1	EO-01-071018-C	4	0.0136	-7470A	PPB	12 Jul 2018 10:40:22	U	SMPL
J3832-02 - 1	TR-05-071018	41	0.0231	-7470A	PPB	12 Jul 2018 10:42:28	U	SMPL
J3918-02 - 1	VNJ-204-255-ETS	93	0.0365	-7470A	PPB	12 Jul 2018 10:44:35	U	SMPL
J3924-02 - 1	RO-1-4	58	0.0275	-7470A	PPB	12 Jul 2018 10:46:41	U	SMPL
J3926-02 - 1	TRAP-BAG	-54	-0.0013	-7470A	PPB	12 Jul 2018 10:48:48	U	SMPL
J3927-04 - 1	CS-RBR-200031	-35	0.0036	-7470A	PPB	12 Jul 2018 10:50:56	U	SMPL
J3928-02 - 1	SOIL-CUTTINGS	-47	0.0005	-7470A	PPB	12 Jul 2018 10:53:04	U	SMPL
J3930-04 - 1	NYS-RBR-200030	-16	0.0085	-7470A	PPB	12 Jul 2018 10:55:12	U	SMPL
J3930-04DUP - 1	NYS-RBR-200030DUP	-12	0.0095	-7470A	PPB	12 Jul 2018 10:57:21	U	SMPL
J3930-04MS - 1	NYS-RBR-200030MS	13075	3.3718	-7470A	PPB	12 Jul 2018 10:59:28	U	SMPL
CCV59 - 1	CCV59	19119	4.9246	-7470A	PPB	12 Jul 2018 11:01:35	U	SMPL
CCB59 - 1	CCB59	-156	-0.0275	-7470A	PPB	12 Jul 2018 11:03:42	U	SMPL
J3930-04MSD - 1	NYS-RBR-200030MSD	13787	3.5547	-7470A	PPB	12 Jul 2018 11:05:50	U	SMPL
PB111015BL - 1	PBW	-119	-0.0180	-7470A	PPB	12 Jul 2018 11:07:57	U	SMPL
PB111015BS - 1	LCSW	13752	3.5457	-7470A	PPB	12 Jul 2018 11:10:04	U	SMPL
J3934-01 - 1	TS-BASELINE-WATER	1367	0.3638	-7470A	PPB	12 Jul 2018 11:12:11	U	SMPL
J3934-01DUP - 1	TS-BASELINE-WATERDUP	1436	0.3815	-7470A	PPB	12 Jul 2018 11:14:19	U	SMPL
J3934-01MS - 1	TS-BASELINE-WATERMS	13464	3.4717	-7470A	PPB	12 Jul 2018 11:16:26	U	SMPL
J3934-01MSD - 1	TS-BASELINE-WATERMSD	13362	3.4455	-7470A	PPB	12 Jul 2018 11:18:33	U	SMPL
CCV60 - 1	CCV60	18628	4.7984	-7470A	PPB	12 Jul 2018 11:20:43	U	SMPL
CCB60 - 1	CCB60	-162	-0.0290	-7470A	PPB	12 Jul 2018 11:22:53	U	SMPL
J3930-04LX5 - 1		-21	0.0072	-7470A	PPB	12 Jul 2018 11:25:01	U	SMPL
J3930-04A - 1		15030	3.8741	-7470A	PPB	12 Jul 2018 11:27:08	U	SMPL
J3934-01LX5 - 1		171	0.0565	-7470A	PPB	12 Jul 2018 11:29:15	U	SMPL
J3934-01A - 1		15106	3.8936	-7470A	PPB	12 Jul 2018 11:31:23	U	SMPL
CCV61 - 1	CCV61	18685	4.8131	-7470A	PPB	12 Jul 2018 11:33:30	U	SMPL
CCB61 - 1	CCB61	-168	-0.0306	-7470A	PPB	12 Jul 2018 11:35:38	U	SMPL