


7471B

LB965-97
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45659



A= 0.0000e+000

B= 2.2009e-004

C= -2.9287e-002

Rho= 0.9999684

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	8	0.000	8				
0.2	0.200	0.227	0.027	1166	0.0 %	1166				
2.5	2.500	2.497	-0.003	11477	0.0 %	11477				
5.0	5.000	5.031	0.031	22993	0.0 %	22993				
7.5	7.500	7.452	-0.048	33993	0.0 %	33993				
10.0	10.000	10.020	0.020	45659	0.0 %	45659				

LB96597 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Jul 2018 12:07:12

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	8	-	0.00007471B	PPB	12 Jul 2018 12:08:21	S	Std
0.2 - 1	502	1166	-	0.20007471B	PPB	12 Jul 2018 12:10:39	S	Std
2.5 - 1	52.5	11477	-	2.50007471B	PPB	12 Jul 2018 12:12:56	S	Std
5.0 - 1	55	22993	-	5.00007471B	PPB	12 Jul 2018 12:15:14	S	Std
7.5 - 1	57.5	33993	-	7.50007471B	PPB	12 Jul 2018 12:17:31	S	Std
10.0 - 1	510	45659	-	10.00007471B	PPB	12 Jul 2018 12:19:48	S	Std
ICV46 - 1	ICV46	18233	3.9836	-7471B	PPB	12 Jul 2018 12:22:34	U	SMPL
ICB46 - 1	ICB46	-119	-0.0555	-7471B	PPB	12 Jul 2018 12:24:50	U	SMPL
CCV62 - 1	CCV62	23686	5.1838	-7471B	PPB	12 Jul 2018 12:27:06	U	SMPL
CCB62 - 1	CCB62	-123	-0.0564	-7471B	PPB	12 Jul 2018 12:29:22	U	SMPL
CRA - 1	CRA	1147	0.2232	-7471B	PPB	12 Jul 2018 12:31:39	U	SMPL
HighStd - 1	HighStd	45797	10.0503	-7471B	PPB	12 Jul 2018 12:33:54	U	SMPL
ChkStd - 1	ChkStd	31445	6.8915	-7471B	PPB	12 Jul 2018 12:36:10	U	SMPL
PB111059BL - 1	PB111059BL	-190	-0.0711	-7471B	PPB	12 Jul 2018 12:38:28	U	SMPL
PB111059BS - 1	PB111059BS	17023	3.7173	-7471B	PPB	12 Jul 2018 12:40:46	U	SMPL
J3633-01 - 1	A-1(0-6)	2056	0.4232	-7471B	PPB	12 Jul 2018 12:43:03	U	SMPL
J3633-03 - 1	B-1(0-6)	1278	0.2520	-7471B	PPB	12 Jul 2018 12:45:20	U	SMPL
J3633-05 - 1	C-1(0-3)	3272	0.6909	-7471B	PPB	12 Jul 2018 12:47:38	U	SMPL
J3825-01 - 1	CL-02-071118-A	5753	1.2369	-7471B	PPB	12 Jul 2018 12:49:55	U	SMPL
J3825-01DUP - 1	CL-02-071118-ADUP	6190	1.3331	-7471B	PPB	12 Jul 2018 12:52:11	U	SMPL
CCV63 - 1	CCV63	23177	5.0718	-7471B	PPB	12 Jul 2018 12:54:27	U	SMPL
CCB63 - 1	CCB63	-149	-0.0621	-7471B	PPB	12 Jul 2018 12:56:43	U	SMPL
J3825-01MS - 1	CL-02-071118-AMS	23205	5.0779	-7471B	PPB	12 Jul 2018 12:59:00	U	SMPL
J3825-01MSD - 1	CL-02-071118-AMSD	21251	4.6479	-7471B	PPB	12 Jul 2018 13:01:15	U	SMPL
J3825-03 - 1	CL-02-071118-B	5797	1.2466	-7471B	PPB	12 Jul 2018 13:03:32	U	SMPL
J3825-05 - 1	CL-02-071118-C	6197	1.3346	-7471B	PPB	12 Jul 2018 13:05:50	U	SMPL
J3825-07 - 1	CL-02-071118-D	6526	1.4070	-7471B	PPB	12 Jul 2018 13:08:06	U	SMPL
J3825-09 - 1	CL-02-071118-E	8311	1.7999	-7471B	PPB	12 Jul 2018 13:10:22	U	SMPL
J3943-01 - 1	C-1	385	0.0554	-7471B	PPB	12 Jul 2018 13:12:39	U	SMPL
J3944-01 - 1	FOUNDRY-WEST	47914	10.5162	-7471B	PPB	12 Jul 2018 13:14:57	U	SMPL
J3944-02 - 1	FOUNDRY-COMP	12621	2.7485	-7471B	PPB	12 Jul 2018 13:17:15	U	SMPL
J3944-03 - 1	FOUNDRY-EAST	2007	0.4124	-7471B	PPB	12 Jul 2018 13:19:33	U	SMPL
CCV64 - 1	CCV64	23491	5.1409	-7471B	PPB	12 Jul 2018 13:21:49	U	SMPL
CCB64 - 1	CCB64	-153	-0.0630	-7471B	PPB	12 Jul 2018 13:24:05	U	SMPL
J3947-01 - 1	PAR-1	1424	0.2841	-7471B	PPB	12 Jul 2018 13:26:22	U	SMPL
J3947-03 - 1	PAR-2	197618	43.4648	-7471B	PPB	12 Jul 2018 13:28:39	U	SMPL
J3947-05 - 1	PAR-3	287514	63.2502	-7471B	PPB	12 Jul 2018 13:30:55	U	SMPL
J3947-07 - 1	PAR-4	274832	60.4590	-7471B	PPB	12 Jul 2018 13:33:25	U	SMPL
J3948-01 - 1	2018-NYSEG-CLARK-ST-PH4-AREA-2-BANK	34420	7.5463	-7471B	PPB	12 Jul 2018 13:36:05	U	SMPL
J3948-03 - 1	2018-NYSEG-CLARK-ST-PH4-AREA-3-BANK	17528	3.8285	-7471B	PPB	12 Jul 2018 13:38:49	U	SMPL
J3948-05 - 1	2018-NYSEG-CLARK-ST-PH4-AREA-5-6-BANK	46918	10.2970	-7471B	PPB	12 Jul 2018 13:41:08	U	SMPL
J3825-01LX5 - 1		909	0.1708	-7471B	PPB	12 Jul 2018 13:46:26	U	SMPL
J3825-01A - 1		23435	5.1286	-7471B	PPB	12 Jul 2018 13:48:43	U	SMPL
CCV65 - 1	CCV65	23328	5.1050	-7471B	PPB	12 Jul 2018 13:50:59	U	SMPL
CCB65 - 1	CCB65	-158	-0.0641	-7471B	PPB	12 Jul 2018 13:53:16	U	SMPL
J3944-01DLX2 - 1		24456	5.3533	-7471B	PPB	12 Jul 2018 13:55:34	U	SMPL
J3947-03DLX10 - 1		20226	4.4223	-7471B	PPB	12 Jul 2018 13:57:50	U	SMPL
J3947-05DLX10 - 1		30843	6.7590	-7471B	PPB	12 Jul 2018 14:00:07	U	SMPL
J3947-07DLX10 - 1		29334	6.4269	-7471B	PPB	12 Jul 2018 14:02:23	U	SMPL
J3948-05DLX2 - 1		23347	5.1092	-7471B	PPB	12 Jul 2018 14:04:40	U	SMPL
CCV66 - 1	CCV66	23168	5.0698	-7471B	PPB	12 Jul 2018 14:06:58	U	SMPL
CCB66 - 1	CCB66	-162	-0.0649	-7471B	PPB	12 Jul 2018 14:09:16	U	SMPL