



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96598

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.6	07/12/2018	12:10
2	J3945-01		1	7	>100.0	>212.0	>212.0	07/12/2018	12:40
3	J3945-01DUP		1	7	>100.0	>212.0	>212.0	07/12/2018	13:10
4	J3947-28		1	6	>100.0	>212.0	>212.0	07/12/2018	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96598

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp71218

WorkList ID : 114385

Date : 7/12/2018 11:41:14 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2018	Water	J3945-01	Flash Point	Cool 4 deg C	DIST01	K52	2-AVE-16	07/11/2018	1010A
07/15/2018	Water	J3947-28	Flash Point	Cool 4 deg C	PSEG01	K41	PAR-1-OILY-WATER	07/11/2018	1010A

Date/Time 7/12/18 11:45AM
 Received by: KS
 Relinquished by: CP

Date/Time 7/12/18 13:45
 Received by: CP
 Relinquished by: KS