

245.1


LB 96610

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35947



A= 0.0000e+000

B= 2.7862e-004

C= -6.7021e-003

Rho= 0.9999835

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	-41	0.000	-41				
0.05	0.050	0.039	-0.011	163	0.0 %	163				
0.2	0.200	0.206	0.006	765	0.0 %	765				
2.5	2.500	2.529	0.029	9102	0.0 %	9102				
5.0	5.000	5.021	0.021	18044	0.0 %	18044				
7.5	7.500	7.464	-0.036	26812	0.0 %	26812				
10.0	10.000	10.009	0.009	35947	0.0 %	35947				

LB96610 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Jul 2018 17:02:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	245.1	PPB	12 Jul 2018 17:04:17	S	Std
0.05 - 1	50.05	163	-	0.0500	245.1	PPB	12 Jul 2018 17:06:25	S	Std
0.2 - 1	50.2	765	-	0.2000	245.1	PPB	12 Jul 2018 17:08:34	S	Std
2.5 - 1	52.5	9102	-	2.5000	245.1	PPB	12 Jul 2018 17:10:42	S	Std
5.0 - 1	55	18044	-	5.0000	245.1	PPB	12 Jul 2018 17:12:50	S	Std
7.5 - 1	57.5	26812	-	7.5000	245.1	PPB	12 Jul 2018 17:14:57	S	Std
10.0 - 1	510	35947	-	10.0000	245.1	PPB	12 Jul 2018 17:19:35	S	Std
ICV47 - 1	ICV47	14181	3.9445	-245.1		PPB	12 Jul 2018 17:27:17	U	SMPL
ICB47 - 1	ICB47	-181	-0.0571	-245.1		PPB	12 Jul 2018 17:29:23	U	SMPL
CCV67 - 1	CCV67	17807	4.9548	-245.1		PPB	12 Jul 2018 17:31:29	U	SMPL
CCB67 - 1	CCB67	-197	-0.0616	-245.1		PPB	12 Jul 2018 17:33:36	U	SMPL
CRA - 1	CRA	165	0.0393	-245.1		PPB	12 Jul 2018 17:35:44	U	SMPL
HighStd - 1	HighStd	36042	10.0355	-245.1		PPB	12 Jul 2018 17:37:51	U	SMPL
ChkStd - 1	ChkStd	24476	6.8129	-245.1		PPB	12 Jul 2018 17:39:58	U	SMPL
PB111016BL - 1	PBW	-241	-0.0739	-245.1		PPB	12 Jul 2018 17:42:07	U	SMPL
PB111016BS - 1	LCSW	13656	3.7982	-245.1		PPB	12 Jul 2018 17:44:15	U	SMPL
J3931-01 - 1	070918-DBRI-E-D-B	-344	-0.1025	-245.1		PPB	12 Jul 2018 17:46:23	U	SMPL
J3931-02 - 1	070918-DBRI-E-D-M	-242	-0.0741	-245.1		PPB	12 Jul 2018 17:48:31	U	SMPL
J3931-03 - 1	070918-DBRI-E-D-S	-227	-0.0699	-245.1		PPB	12 Jul 2018 17:50:39	U	SMPL
J3931-04 - 1	070918-DBRI-E-U-B	-176	-0.0557	-245.1		PPB	12 Jul 2018 17:52:47	U	SMPL
J3931-05 - 1	070918-DBRI-E-U-M	-226	-0.0697	-245.1		PPB	12 Jul 2018 17:54:54	U	SMPL
CCV68 - 1	CCV68	17863	4.9704	-245.1		PPB	12 Jul 2018 17:57:01	U	SMPL
CCB68 - 1	CCB68	-194	-0.0608	-245.1		PPB	12 Jul 2018 17:59:07	U	SMPL
J3931-06 - 1	070918-DBRI-E-U-S	-265	-0.0805	-245.1		PPB	12 Jul 2018 18:01:15	U	SMPL
J3931-06DUP - 1	070918-DBRI-E-U-SDUP	-224	-0.0691	-245.1		PPB	12 Jul 2018 18:03:22	U	SMPL
J3931-06MS - 1	070918-DBRI-E-U-SMS	13654	3.7976	-245.1		PPB	12 Jul 2018 18:05:29	U	SMPL
J3931-06MSD - 1	070918-DBRI-E-U-SMSD	12831	3.5683	-245.1		PPB	12 Jul 2018 18:07:36	U	SMPL
J3931-06LX5 - 1		-194	-0.0608	-245.1		PPB	12 Jul 2018 18:09:43	U	SMPL
J3931-06A - 1		15937	4.4337	-245.1		PPB	12 Jul 2018 18:11:51	U	SMPL
CCV69 - 1	CCV69	17712	4.9283	-245.1		PPB	12 Jul 2018 18:13:59	U	SMPL
CCB69 - 1	CCB69	-205	-0.0638	-245.1		PPB	12 Jul 2018 18:16:08	U	SMPL