



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96612

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/12/2018	16:40
2	CAL2	1	Water			21.5	7.00	07/12/2018	16:42
3	CAL3	1	Water			21.8	10.03	07/12/2018	16:44
4	ICV	1	Water			21.7	7.00	07/12/2018	16:46
5	CCV1	1	Water			21.7	2.00	07/12/2018	16:48
6	J3976-01	1	Solid	20.02	20	24.7	6.34	07/12/2018	16:50
7	J3976-01DUP	1	Solid	20.03	20	24.8	6.35	07/12/2018	16:52
8	CCV2	1	Water			21.6	12.00	07/12/2018	16:54

WORKLIST(Hardcopy Internal Chain)

LP 96612

WorkList Name : pH071218-1

WorkList ID : 114410

Date : 7/12/2018 4:29:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2018	Solid	J3976-01	pH	Cool 4 deg C	TULL02	K61	FK-PS-01-033	07/12/2018	9045C

Date/Time 07/12/18 4:35 PM

Received by: SP

Relinquished by: CP

Date/Time 07/12/18 5:36

Received by: CP

Relinquished by: SP