



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96614

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/12/2018	16:40
2	CAL2	1	Water			21.5	7.00	07/12/2018	16:42
3	CAL3	1	Water			21.8	10.03	07/12/2018	16:44
4	ICV	1	Water			21.7	7.00	07/12/2018	16:46
5	CCV1	1	Water			21.6	2.00	07/12/2018	17:05
6	J3545-05	1	Water			22.0	5.49	07/12/2018	17:07
7	J3545-05DUP	1	Water			22.1	5.50	07/12/2018	17:09
8	J3954-01	1	Water			23.6	11.07	07/12/2018	17:11
9	J3964-01	1	Water			24.6	6.54	07/12/2018	17:15
10	J3975-13	1	Water			24.3	6.95	07/12/2018	17:17
11	CCV2	1	Water			21.7	12.00	07/12/2018	17:19

WORKLIST(Hardcopy Internal Chain)

LB 96614

WorkList Name : pH071218

WorkList ID : 114399

Date : 7/12/2018 2:01:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2018	Water	J3545-05	pH	Cool 4 deg C	PSEG01	K61	DRUM-1	07/12/2018	9040C
07/12/2018	Water	J3954-01	pH	Cool 4 deg C	PSEG01	K41	FM1041	07/12/2018	9040C
07/12/2018	Water	J3964-01	pH	Cool 4 deg C	PSEG01	K42	N48967	07/12/2018	9040C
07/12/2018	Water	J3975-13	pH	Cool 4 deg C	PSEG01	K51	SV30518L	07/12/2018	9040C

Date/Time 07/12/18 5:00 PM

Received by: SP

Relinquished by: CP

Date/Time 07/12/18 5:30

Received by: CP

Relinquished by: SP