



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 07/13/2018TEMP1 IN: 103 07/12/2018 16:00 TEMP OUT: 104 07/13/2018 07:15TEMP2 IN: 104 07/13/2018 07:50 TEMP OUT: 104 07/13/2018 09:00TEMP3 IN: 105 07/13/2018 11:00 TEMP OUT: 105 07/13/2018 12:35TEMP4 IN: 105 07/13/2018 13:10 TEMP OUT: 105 07/13/2018 14:35Run Number: LB96622BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96622BL	LB96622BL	1.4135	1.4135	100	1.4135	1.4135	1.4135	0.0000	0
2	LB96622BS	LB96622BS	1.3943	1.3943	100	1.4450	1.4450	1.4450	0.0507	507
3	J3795-01	USD-MF-01	1.3987	1.3987	400	1.4036	1.4036	1.4036	0.0049	12.3
4	J3969-01	UDS-CH-4-01	1.4052	1.4052	100	1.4110	1.4110	1.4110	0.0058	58
5	J3969-01DUP	UDS-CH-4-01DUP	1.4159	1.4159	100	1.4215	1.4215	1.4215	0.0056	56
6	J3969-02	UDS-CH-22-01	1.3989	1.3989	100	1.4018	1.4018	1.4018	0.0029	29
7	J3979-01	001M-WILLETS-PT-BLVD (JULY)	1.4013	1.4013	150	1.4079	1.4079	1.4079	0.0066	44
8	J3979-02	002M-35TH-AVE (JULY)	1.4142	1.4142	150	1.4219	1.4219	1.4219	0.0077	51.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS071318-2

WorkList ID : 114415

Date : 7/13/2018 7:30:22 AM

2096622

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/13/2018	Water	J3795-01	N/D TSS	Cool 4 deg C	LIRO01	K13	USD-MF-01	07/11/2018	SM2540 D
07/13/2018	Water	J3969-01	N/D TSS	Cool 4 deg C	LIRO01	L12	UDS-CH-4-01	07/11/2018	SM2540 D
07/13/2018	Water	J3969-02	N/D TSS	Cool 4 deg C	LIRO01	L12	UDS-CH-22-01	07/11/2018	SM2540 D
07/15/2018	Water	J3979-01	N TSS	Cool 4 deg C	TULL01	B22	001M-WILLETS-PT-BLVD(JUL	07/11/2018	SM2540 D
07/15/2018	Water	J3979-02	N TSS	Cool 4 deg C	TULL01	B22	002M-35TH-AVE(JULY)	07/11/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

07/13/18, 08:00
AS
CS

Date/Time

Received by:

Relinquished by:

07/13/18 12:30
CS
AS