

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB96632

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP66083
calibration std. phosphate 0.5 ppm	WP66082
calibration std. phosphate 0.3 ppm	WP66081
calibration std. 0 ppm	WP66078
phosphate ICV std.	WP66086
phosphate CCV std.	WP66087
calibration std. phosphate 0.1 ppm	WP66080
calibration std. phosphate 0.05 ppm	WP66079
Combined reagent	WP66119
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.001Slope: 0.6502Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/13/2018	11:00
2	CAL2	0.05	1	50	50	0.033	07/13/2018	11:01
3	CAL3	0.10	1	50	50	0.065	07/13/2018	11:01
4	CAL4	0.30	1	50	50	0.186	07/13/2018	11:02
5	CAL5	0.50	1	50	50	0.329	07/13/2018	11:02
6	CAL6	1.00	1	50	50	0.649	07/13/2018	11:04



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB96632

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.500	07/13/2018	11:05
2	ICB		1	50	50	0.000	0.002	07/13/2018	11:07
3	CCV1	0.50	1	50	50	0.328	0.506	07/13/2018	11:07
4	CCB1		1	50	50	0.000	0.002	07/13/2018	11:08
5	PB111056BL		1	50	50	0.000	0.002	07/13/2018	11:09
6	PB111056BS	0.50	1	50	50	0.321	0.495	07/13/2018	11:09
7	J3941-01		1	50	50	0.572	0.881	07/13/2018	11:10
8	J3941-01DUP		1	50	50	0.560	0.863	07/13/2018	11:10
9	J3941-01MS	0.50	1	50	50	0.722	1.112	07/13/2018	11:11
10	J3941-01MSD	0.50	1	50	50	0.749	1.153	07/13/2018	11:12
11	J3968-01		1	50	50	0.429	0.661	07/13/2018	11:13
12	J3941-01MS	0.50	2	50	50	0.481	0.741	07/13/2018	11:29
13	J3941-01MSD	0.50	2	50	50	0.475	0.732	07/13/2018	11:30
14	CCV2	0.50	1	50	50	0.326	0.503	07/13/2018	11:30
15	CCB2		1	50	50	0.000	0.002	07/13/2018	11:31