

LB 96638


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

49391



A= 0.0000e+000

B= 2.0384e-004

C= -4.4993e-002

Rho= 0.9999424

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	348	0.000	348				
0.2	0.200	0.203	0.003	1217	0.0 %	1217				
2.5	2.500	2.444	-0.056	12209	0.0 %	12209				
5.0	5.000	5.052	0.052	25007	0.0 %	25007				
7.5	7.500	7.452	-0.048	36780	0.0 %	36780				
10.0	10.000	10.023	0.023	49391	0.0 %	49391				

LB96638 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jul 2018 15:38:05

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	348	-	0.0000	7471B	PPB	13 Jul 2018 15:43:43	S	Std
0.2 - 1	50.2	1217	-	0.2000	7471B	PPB	13 Jul 2018 15:46:00	S	Std
2.5 - 1	52.5	12209	-	2.5000	7471B	PPB	13 Jul 2018 15:48:17	S	Std
5.0 - 1	55	25007	-	5.0000	7471B	PPB	13 Jul 2018 15:50:35	S	Std
7.5 - 1	57.5	36780	-	7.5000	7471B	PPB	13 Jul 2018 15:55:23	S	Std
10.0 - 1	510	49391	-	10.0000	7471B	PPB	13 Jul 2018 16:00:16	S	Std
ICV49 - 1	ICV49	18858	3.7990	-	7471B	PPB	13 Jul 2018 16:02:56	U	SMPL
ICB49 - 1	ICB49	160	-0.0124	-	7471B	PPB	13 Jul 2018 16:05:12	U	SMPL
CCV79 - 1	CCV79	23848	4.8161	-	7471B	PPB	13 Jul 2018 16:07:28	U	SMPL
CCB79 - 1	CCB79	156	-0.0132	-	7471B	PPB	13 Jul 2018 16:09:44	U	SMPL
CRA - 1	CRA	1211	0.2019	-	7471B	PPB	13 Jul 2018 16:12:00	U	SMPL
HighStd - 1	HighStd	50012	10.1493	-	7471B	PPB	13 Jul 2018 16:14:15	U	SMPL
ChkStd - 1	ChkStd	32730	6.6266	-	7471B	PPB	13 Jul 2018 16:16:32	U	SMPL
PB111082BL - 1	PBS	-15	-0.0481	-	7471B	PPB	13 Jul 2018 16:18:50	U	SMPL
PB111082BS - 1	LCSS	17390	3.4997	-	7471B	PPB	13 Jul 2018 16:21:08	U	SMPL
J3542-01 - 1	LL-SOIL-1	6477	1.2753	-	7471B	PPB	13 Jul 2018 16:23:25	U	SMPL
J3542-01DUP - 1	LL-SOIL-1DUP	5848	1.1471	-	7471B	PPB	13 Jul 2018 16:25:42	U	SMPL
J3542-01MS - 1	LL-SOIL-1MS	24422	4.9331	-	7471B	PPB	13 Jul 2018 16:33:13	U	SMPL
J3542-01MSD - 1	LL-SOIL-1MSD	24943	5.0393	-	7471B	PPB	13 Jul 2018 16:35:30	U	SMPL
J3542-05 - 1	LL-SOIL-2	6258	1.2306	-	7471B	PPB	13 Jul 2018 16:37:47	U	SMPL
CCV80 - 1	CCV80	22889	4.6207	-	7471B	PPB	13 Jul 2018 16:40:04	U	SMPL
CCB80 - 1	CCB80	147	-0.0150	-	7471B	PPB	13 Jul 2018 16:42:20	U	SMPL
J3542-09 - 1	LL-SOIL-3	5294	1.0341	-	7471B	PPB	13 Jul 2018 16:44:37	U	SMPL
J3545-02 - 1	DRUM-3	111	-0.0224	-	7471B	PPB	13 Jul 2018 16:46:52	U	SMPL
J3820-03 - 1	BU-02-071218	2497	0.4640	-	7471B	PPB	13 Jul 2018 16:49:08	U	SMPL
J3821-01 - 1	HR-05-071218-A	6572	1.2946	-	7471B	PPB	13 Jul 2018 16:51:24	U	SMPL
J3821-03 - 1	HR-05-071218-B	8895	1.7681	-	7471B	PPB	13 Jul 2018 16:53:41	U	SMPL
J3821-05 - 1	HR-05-071218-C	6627	1.3058	-	7471B	PPB	13 Jul 2018 16:55:58	U	SMPL
J3831-01 - 1	SU-04-071218	1248	0.2094	-	7471B	PPB	13 Jul 2018 16:58:15	U	SMPL
J3956-01 - 1	CONTINGENCY-DUCK-BANK-CONCRETE-A	336	0.0235	-	7471B	PPB	13 Jul 2018 17:00:32	U	SMPL
J3956-03 - 1	CONTINGENCY-DUCK-BANK-CONCRETE-B	349	0.0261	-	7471B	PPB	13 Jul 2018 17:02:50	U	SMPL
PB111083BL - 1	PBS	164	-0.0116	-	7471B	PPB	13 Jul 2018 17:05:06	U	SMPL
CCV81 - 1	CCV81	24030	4.8532	-	7471B	PPB	13 Jul 2018 17:07:22	U	SMPL
CCB81 - 1	CCB81	67	-0.0313	-	7471B	PPB	13 Jul 2018 17:09:39	U	SMPL
PB111083BS - 1	LCSS	16919	3.4037	-	7471B	PPB	13 Jul 2018 17:11:55	U	SMPL
J3958-01 - 1	CS-STONE	504	0.0577	-	7471B	PPB	13 Jul 2018 17:14:11	U	SMPL
J3958-07 - 1	CS-CONCRETE	572	0.0716	-	7471B	PPB	13 Jul 2018 17:16:28	U	SMPL
J3966-01 - 1	72-11977	1259	0.2116	-	7471B	PPB	13 Jul 2018 17:18:44	U	SMPL
J3966-01DUP - 1	72-11977DUP	1206	0.2008	-	7471B	PPB	13 Jul 2018 17:24:14	U	SMPL
J3966-01MS - 1	72-11977MS	19024	3.8328	-	7471B	PPB	13 Jul 2018 17:29:34	U	SMPL
J3966-01MSD - 1	72-11977MSD	19329	3.8950	-	7471B	PPB	13 Jul 2018 17:31:51	U	SMPL
J3975-01 - 1	PAR-5	41622	8.4391	-	7471B	PPB	13 Jul 2018 17:34:09	U	SMPL
J3975-03 - 1	PAR-6	97101	19.7479	-	7471B	PPB	13 Jul 2018 17:36:26	U	SMPL
J3975-05 - 1	PAR-7	5234	1.0219	-	7471B	PPB	13 Jul 2018 17:38:44	U	SMPL
CCV82 - 1	CCV82	23949	4.8367	-	7471B	PPB	13 Jul 2018 17:41:04	U	SMPL
CCB82 - 1	CCB82	-55	-0.0562	-	7471B	PPB	13 Jul 2018 17:43:20	U	SMPL
J3975-07 - 1	PAR-8	67732	13.7614	-	7471B	PPB	13 Jul 2018 17:45:37	U	SMPL
J3976-01 - 1	FK-PS-01-033	1049	0.1688	-	7471B	PPB	13 Jul 2018 17:47:53	U	SMPL
J3980-03 - 1	TP-E-MH-S	3266	0.6207	-	7471B	PPB	13 Jul 2018 17:50:12	U	SMPL
J3980-07 - 1	TP-E-MH-D	833	0.1248	-	7471B	PPB	13 Jul 2018 17:52:28	U	SMPL
J3542-01LX5 - 1		1320	0.2241	-	7471B	PPB	13 Jul 2018 17:54:45	U	SMPL
J3542-01A - 1		24243	4.8967	-	7471B	PPB	13 Jul 2018 17:57:01	U	SMPL
J3966-01LX5 - 1		184	-0.0075	-	7471B	PPB	13 Jul 2018 17:59:18	U	SMPL
J3966-01A - 1		17872	3.5980	-	7471B	PPB	13 Jul 2018 18:01:36	U	SMPL
J3975-03DLX5 - 1		20477	4.1290	-	7471B	PPB	13 Jul 2018 18:03:54	U	SMPL
J3975-07DLX5 - 1		13836	2.7753	-	7471B	PPB	13 Jul 2018 18:06:10	U	SMPL
CCV83 - 1	CCV83	24219	4.8918	-	7471B	PPB	13 Jul 2018 18:08:27	U	SMPL
CCB83 - 1	CCB83	-18	-0.0487	-	7471B	PPB	13 Jul 2018 18:10:43	U	SMPL