

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/16/2018

TEMP1 IN: 103 07/16/2018 07:10 TEMP OUT: 104 07/16/2018 08:15

TEMP2 IN: 104 07/16/2018 09:00 TEMP OUT: 104 07/16/2018 10:05

TEMP3 IN: 105 07/16/2018 11:15 TEMP OUT: 105 07/16/2018 12:50

TEMP4 IN: 105 07/16/2018 13:45 TEMP OUT: 105 07/16/2018 15:30

Run Number: LB96654

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96654BL	LB96654BL	1.3952	1.3952	100	1.3952	1.3952	1.3952	0.0000	0
2	LB96654BS	LB96654BS	1.4050	1.4050	100	1.4550	1.4550	1.4550	0.0500	500
3	J3990-01	071218-DBRI-F-D-B	1.4132	1.4132	200	1.4329	1.4329	1.4329	0.0197	98.5
4	J3990-01DUP	071218-DBRI-F-D-BDUP	1.4053	1.4053	200	1.4260	1.4260	1.4260	0.0207	103.5
5	J3990-02	071218-DBRI-F-D-M	1.3991	1.3991	300	1.4147	1.4147	1.4147	0.0156	52
6	J3990-03	071218-DBRI-F-D-S	1.4011	1.4011	400	1.4143	1.4143	1.4143	0.0132	33
7	J3990-04	071218-DBRI-F-U-B	1.4022	1.4022	200	1.4275	1.4275	1.4275	0.0253	126.5
8	J3990-05	071218-DBRI-F-U-M	1.3966	1.3966	200	1.4141	1.4141	1.4141	0.0175	87.5
9	J3990-06	071218-DBRI-F-U-S	1.4125	1.4125	400	1.4189	1.4189	1.4189	0.0064	16
10	J3990-07	071218-TIPC-F-D-B	1.4052	1.4052	300	1.4200	1.4200	1.4200	0.0148	49.3
11	J3990-08	071218-TIPC-F-D-S	1.4071	1.4071	300	1.4194	1.4194	1.4194	0.0123	41
12	J3990-09	071218-TIPC-F-U-B	1.4003	1.4003	300	1.4141	1.4141	1.4141	0.0138	46
13	J3990-10	071218-TIPC-F-U-S	1.4076	1.4076	400	1.4227	1.4227	1.4227	0.0151	37.8



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 07/16/2018

TEMP1 IN: 103 07/16/2018 07:10 TEMP OUT: 104 07/16/2018 08:15

TEMP2 IN: 104 07/16/2018 09:00 TEMP OUT: 104 07/16/2018 10:05

TEMP3 IN: 105 07/16/2018 11:15 TEMP OUT: 105 07/16/2018 12:50

TEMP4 IN: 105 07/16/2018 13:45 TEMP OUT: 105 07/16/2018 15:30

Run Number: LB96654

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS071618

WorkList ID : 114479

Date : 7/16/2018 7:11:38 AM

21296654

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2018	Water	J3990-01 C	TSS	Cool 4 deg C	HDR103	B32	071218-DBRI-F-D-B	07/12/2018	SM2540 D
07/19/2018	Water	J3990-02 C	TSS	Cool 4 deg C	HDR103	B32	071218-DBRI-F-D-M	07/12/2018	SM2540 D
07/19/2018	Water	J3990-03 C	TSS	Cool 4 deg C	HDR103	B32	071218-DBRI-F-D-S	07/12/2018	SM2540 D
07/19/2018	Water	J3990-04 C	TSS	Cool 4 deg C	HDR103	B32	071218-DBRI-F-U-B	07/12/2018	SM2540 D
07/19/2018	Water	J3990-05 C	TSS	Cool 4 deg C	HDR103	B32	071218-DBRI-F-U-M	07/12/2018	SM2540 D
07/19/2018	Water	J3990-06 C	TSS	Cool 4 deg C	HDR103	B32	071218-DBRI-F-U-S	07/12/2018	SM2540 D
07/19/2018	Water	J3990-07 C	TSS	Cool 4 deg C	HDR103	B32	071218-TIPC-F-D-B	07/12/2018	SM2540 D
07/19/2018	Water	J3990-08 C	TSS	Cool 4 deg C	HDR103	B32	071218-TIPC-F-D-S	07/12/2018	SM2540 D
07/19/2018	Water	J3990-09 C	TSS	Cool 4 deg C	HDR103	B32	071218-TIPC-F-U-B	07/12/2018	SM2540 D
07/19/2018	Water	J3990-10 C	TSS	Cool 4 deg C	HDR103	B32	071218-TIPC-F-U-S	07/12/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

07/16/18, 08:00

AS

CS

Date/Time

Received by:

Relinquished by:

07/16/18, 12:00

CS

AS