



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96660

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 6                 | 27.0       | 80.6      | 80.6            | 07/16/2018 | 09:40     |
| 2   | J3545-06    |               | 1  | 5                 | >100.0     | >212.0    | >212.0          | 07/16/2018 | 10:02     |
| 3   | J3954-02    |               | 1  | 7                 | >100.0     | >212.0    | >212.0          | 07/16/2018 | 10:42     |
| 4   | J3954-02DUP |               | 1  | 7                 | >100.0     | >212.0    | >212.0          | 07/16/2018 | 11:20     |
| 5   | J3959-02    |               | 1  | 8                 | >100.0     | >212.0    | >212.0          | 07/16/2018 | 11:45     |
| 6   | J3964-02    |               | 1  | 6                 | >100.0     | >212.0    | >212.0          | 07/16/2018 | 12:15     |
| 7   | J3975-14    |               | 1  | 7                 | >100.0     | >212.0    | >212.0          | 07/16/2018 | 12:45     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB96660

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71618

WorkList ID : 114499

Date : 7/16/2018 9:03:05 AM

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 07/17/2018 | Water  | J3545-06 | Flash Point | Cool 4 deg C | PSEG01   | K61              | DRUM-1          | 07/12/2018   | 1010A  |
| 07/17/2018 | Water  | J3954-02 | Flash Point | Cool 4 deg C | PSEG01   | K41              | FM1041          | 07/12/2018   | 1010A  |
| 07/26/2018 | Water  | J3959-02 | Flash Point | Cool 4 deg C | PARS02   | K32              | WC-20180712     | 07/12/2018   | 1010A  |
| 07/17/2018 | Water  | J3964-02 | Flash Point | Cool 4 deg C | PSEG01   | K42              | N48967          | 07/12/2018   | 1010A  |
| 07/17/2018 | Water  | J3975-14 | Flash Point | Cool 4 deg C | PSEG01   | K51              | SV30518L        | 07/12/2018   | 1010A  |

Date/Time 7/16/18 9:10 AM  
 Received by: KS  
 Relinquished by: CP

Date/Time 7/16/18 12:50 PM  
 Received by: CP  
 Relinquished by: KS