

LB 96666


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

45979



A= 0.0000e+000

B= 2.1675e-004

C= -2.3604e-002

Rho= 0.9998381

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	245	0.000	245				
0.2	0.200	0.170	-0.030	892	0.0 %	892				
2.5	2.500	2.516	0.016	11717	0.0 %	11717				
5.0	5.000	4.921	-0.079	22811	0.0 %	22811				
7.5	7.500	7.622	0.122	35272	0.0 %	35272				
10.0	10.000	9.942	-0.058	45979	0.0 %	45979				

LB96666 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jul 2018 14:11:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	245	-	0.0007471B	-	PPB	16 Jul 2018 14:12:41	S	Std
0.2 - 1	50.2	892	-	0.20007471B	-	PPB	16 Jul 2018 14:14:58	S	Std
2.5 - 1	52.5	11717	-	2.50007471B	-	PPB	16 Jul 2018 14:17:16	S	Std
5.0 - 1	55	22811	-	5.00007471B	-	PPB	16 Jul 2018 14:19:33	S	Std
7.5 - 1	57.5	35272	-	7.50007471B	-	PPB	16 Jul 2018 14:21:50	S	Std
10.0 - 1	51.0	45979	-	10.00007471B	-	PPB	16 Jul 2018 14:24:08	S	Std
ICV52 - 1	ICV52	17135	3.6904	-7471B	-	PPB	16 Jul 2018 14:26:53	U	SMPL
ICB52 - 1	ICB52	209	0.0217	-7471B	-	PPB	16 Jul 2018 14:29:08	U	SMPL
CCV96 - 1	CCV96	23033	4.9688	-7471B	-	PPB	16 Jul 2018 14:31:24	U	SMPL
CCB96 - 1	CCB96	210	0.0219	-7471B	-	PPB	16 Jul 2018 14:33:39	U	SMPL
CRA - 1	CRA	891	0.1695	-7471B	-	PPB	16 Jul 2018 14:35:56	U	SMPL
HighStd - 1	HighStd	46800	10.1203	-7471B	-	PPB	16 Jul 2018 14:38:12	U	SMPL
ChkStd - 1	ChkStd	31836	6.8769	-7471B	-	PPB	16 Jul 2018 14:40:28	U	SMPL
PB111156BL - 1	PBS	-302	-0.0891	-7471B	-	PPB	16 Jul 2018 14:42:46	U	SMPL
PB111156BS - 1	LCSS	15141	3.2582	-7471B	-	PPB	16 Jul 2018 14:45:04	U	SMPL
J3816-01 - 1	159-CONC-FLOOR	24862	5.3653	-7471B	-	PPB	16 Jul 2018 14:47:56	U	SMPL
J3816-02 - 1	159-WHITE-BLOCK	350485	75.9443	-7471B	-	PPB	16 Jul 2018 14:50:12	U	SMPL
J3816-03 - 1	159-WHITE-BRICK	230218	49.8763	-7471B	-	PPB	16 Jul 2018 14:52:32	U	SMPL
J3816-04 - 1	159-WHITE-BLUE-BRICK	306664	HIGH	-7471B	-	PPB	16 Jul 2018 14:55:16	X	
J3816-05 - 1	159-BLUE-BRICK	62091	13.4347	-7471B	-	PPB	16 Jul 2018 14:57:30	U	SMPL
CCV97 - 1	CCV97	22639	4.8834	-7471B	-	PPB	16 Jul 2018 14:59:56	U	SMPL
CCB97 - 1	CCB97	-39	-0.0321	-7471B	-	PPB	16 Jul 2018 15:02:15	U	SMPL
J3816-06 - 1	161-TAN-BRICK	862	0.1632	-7471B	-	PPB	16 Jul 2018 15:04:32	U	SMPL
J3816-09 - 1	161-CONC-GREEN-WALL	1314	0.2612	-7471B	-	PPB	16 Jul 2018 15:06:48	U	SMPL
J3816-09DUP - 1	161-CONC-GREEN-WALLDUP	1178	0.2317	-7471B	-	PPB	16 Jul 2018 15:09:04	U	SMPL
J3816-09MS - 1	161-CONC-GREEN-WALLMS	17747	3.8231	-7471B	-	PPB	16 Jul 2018 15:11:20	U	SMPL
J3816-09MSD - 1	161-CONC-GREEN-WALLMSD	17673	3.8070	-7471B	-	PPB	16 Jul 2018 15:13:36	U	SMPL
J3816-09LX5 - 1		35	-0.0160	-7471B	-	PPB	16 Jul 2018 15:15:53	U	SMPL
J3816-09A - 1		17759	3.8257	-7471B	-	PPB	16 Jul 2018 15:18:11	U	SMPL
J3816-02DLX10 - 1		42984	9.2932	-7471B	-	PPB	16 Jul 2018 15:20:28	U	SMPL
J3816-03DLX10 - 1		23781	5.1309	-7471B	-	PPB	16 Jul 2018 15:22:46	U	SMPL
J3816-04DLX10 - 1		42076	9.0964	-7471B	-	PPB	16 Jul 2018 15:25:03	U	SMPL
CCV98 - 1	CCV98	23037	4.9697	-7471B	-	PPB	16 Jul 2018 15:27:20	U	SMPL
CCB98 - 1	CCB98	93	-0.0034	-7471B	-	PPB	16 Jul 2018 15:29:38	U	SMPL
J3816-05DLX2 - 1		32628	7.0485	-7471B	-	PPB	16 Jul 2018 15:31:55	U	SMPL
CCV99 - 1	CCV99	22153	4.7781	-7471B	-	PPB	16 Jul 2018 15:34:11	U	SMPL
CCB99 - 1	CCB99	102	-0.0015	-7471B	-	PPB	16 Jul 2018 15:36:28	U	SMPL