



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96676

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/16/2018	16:10
2	CAL2	1	Water			21.8	7.00	07/16/2018	16:12
3	CAL3	1	Water			21.5	10.03	07/16/2018	16:14
4	ICV	1	Water			21.7	7.00	07/16/2018	16:16
5	CCV1	1	Water			21.7	2.00	07/16/2018	16:18
6	J4006-02	1	Water			25.2	7.39	07/16/2018	16:20
7	J4006-02DUP	1	Water			25.3	7.40	07/16/2018	16:22
8	J4008-02	1	Water			25.5	7.45	07/16/2018	16:24
9	CCV2	1	Water			21.6	12.00	07/16/2018	16:26

WORKLIST(Hardcopy Internal Chain)

LB96676

WorkList Name : corrosivity071618-1

WorkList ID : 114525

Date : 7/16/2018 3:43:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2018	Water	J4006-02	Corrosivity	Cool 4 deg C	PSEG01	K62	RR-BASELINE-WATER	07/16/2018	9040C
07/16/2018	Water	J4008-02	Corrosivity	Cool 4 deg C	PSEG01	K61	KG-FM1026	07/16/2018	9040C

Date/Time 07/16/18 4:00PM

Received by: SP

Relinquished by: CP

Date/Time 07/16/18 4:30

Received by: CP

Relinquished by: SP