



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96679

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	07/16/2018	16:10
2	CAL2	1	Water			21.8	7.00	07/16/2018	16:12
3	CAL3	1	Water			21.5	10.03	07/16/2018	16:14
4	ICV	1	Water			21.7	7.00	07/16/2018	16:16
5	CCV1	1	Water			21.7	2.00	07/16/2018	16:40
6	J3826-04	1	Solid	20.02	20	24.6	8.90	07/16/2018	16:42
7	J3826-04DUP	1	Solid	20.02	20	24.7	8.91	07/16/2018	16:44
8	J3832-04	1	Solid	20.04	20	24.2	8.38	07/16/2018	16:46
9	J4003-01	1	Solid	20.03	20	24.1	6.58	07/16/2018	16:48
10	J4012-03	1	Solid	20.02	20	24.7	8.20	07/16/2018	16:50
11	J4013-02	1	Solid	20.03	20	23.6	6.55	07/16/2018	16:52
12	J4013-06	1	Solid	20.02	20	23.7	6.15	07/16/2018	16:54
13	CCV2	1	Water			21.6	12.00	07/16/2018	16:56

WORKLIST(Hardcopy Internal Chain)

LB 96679

WorkList Name : corrosivity071618

WorkList ID : 114527

Date : 7/16/2018 3:52:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2018	Solid	J3826-04	Corrosivity	Cool 4 deg C	PSEG05		OK-01-071618	07/16/2018	9045C
07/16/2018	Solid	J3832-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-071618	07/16/2018	9045C
07/16/2018	Solid	J4003-01	Corrosivity	Cool 4 deg C	PSEG01	K61	DRUM-1	07/16/2018	9045C
07/16/2018	Solid	J4012-03	Corrosivity	Cool 4 deg C	PSEG01	K63	FS-RT-4917	07/16/2018	9045C
07/16/2018	Solid	J4013-02	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-I-S	07/16/2018	9045C
07/16/2018	Solid	J4013-06	Corrosivity	Cool 4 deg C	PSEG01	K63	TP-I-D	07/16/2018	9045C

Date/Time 07/16/18 4:30 PM

Received by: SP

Relinquished by: SP

Date/Time 07/16/18 5:16 PM

Received by: SP

Relinquished by: SP