


ISM02.4

LB 96685
INSTRUMENT ID: CV1 μ Abs.:

44133



Linear

A= 0.0000e+000

B= 2.2718e-004

C= -6.8634e-004

Rho= 0.9999278

Accept=Accepted

slope
 $y = 1.31e5 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	12	0.000	12					1
0.2	0.200	0.223	0.023	984	0.0 %	984					12
2.5	2.500	2.527	0.027	11128	0.0 %	11128					1
5.0	5.000	4.903	-0.097	21583	0.0 %	21583					-2
7.5	7.500	7.519	0.019	33101	0.0 %	33101					0
10.0	10.000	10.026	0.026	44133	0.0 %	44133					0

AB
7/17/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	12	-		0 ISM02.4	PPB	7/17/2018 9:20	Std
	0.2 S0.2	984	-		0.2 ISM02.4	PPB	7/17/2018 9:22	Std
	2.5 S2.5	11128	-		2.5 ISM02.4	PPB	7/17/2018 9:24	Std
	5 S5	21583	-		5 ISM02.4	PPB	7/17/2018 9:26	Std
	7.5 S7.5	33101	-		7.5 ISM02.4	PPB	7/17/2018 9:29	Std
	10 S10.0	44133	-		10 ISM02.4	PPB	7/17/2018 9:31	Std
ICV	ICV	16835	3.824 -		ISM02.4	PPB	7/17/2018 9:34	SMPL
ICB	ICB	-125	-0.0291 -		ISM02.4	PPB	7/17/2018 9:36	SMPL
CCV006	CCV006	21449	4.8722 -		ISM02.4	PPB	7/17/2018 9:38	SMPL
CCB006	CCB006	-127	-0.0295 -		ISM02.4	PPB	7/17/2018 9:41	SMPL
PB111153BL	PBW010	164	0.0366 -		ISM02.4	PPB	7/17/2018 9:43	SMPL
PB111153TB	LEB010	173	0.0386 -		ISM02.4	PPB	7/17/2018 9:45	SMPL
J3940-01	MBEMG7	-53	-0.0127 -		ISM02.4	PPB	7/17/2018 9:47	SMPL
J3940-02	MBEMG8	-61	-0.0145 -		ISM02.4	PPB	7/17/2018 9:50	SMPL
J3940-03	MBEMG9	-19	-0.005 -		ISM02.4	PPB	7/17/2018 9:52	SMPL
J3940-04	MBEMH0	-177	-0.0409 -		ISM02.4	PPB	7/17/2018 9:54	SMPL
J3940-05	MBEMH0D	9	0.0014 -		ISM02.4	PPB	7/17/2018 9:56	SMPL
J3940-06	MBEMH0S	5006	1.1366 -		ISM02.4	PPB	7/17/2018 9:59	SMPL
CCV007	CCV007	21455	4.8736 -		ISM02.4	PPB	7/17/2018 10:01	SMPL
CCB007	CCB007	-121	-0.0282 -		ISM02.4	PPB	7/17/2018 10:03	SMPL