



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96687

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/17/2018	11:00
2	CAL2	1	Water			21.8	7.00	07/17/2018	11:02
3	CAL3	1	Water			21.7	10.03	07/17/2018	11:04
4	ICV	1	Water			21.7	7.00	07/17/2018	11:06
5	CCV1	1	Water			21.9	2.00	07/17/2018	11:08
6	J4016-03	1	Solid	20.02	20	24.3	8.69	07/17/2018	11:10
7	J4016-03DUP	1	Solid	20.02	20	24.4	8.70	07/17/2018	11:12
8	J4016-06	1	Solid	20.03	20	23.6	8.77	07/17/2018	11:14
9	J4016-09	1	Solid	20.02	20	23.4	9.14	07/17/2018	11:16
10	J4016-12	1	Solid	20.04	20	23.4	9.24	07/17/2018	11:18
11	J4016-15	1	Solid	20.02	20	23.3	9.20	07/17/2018	11:20
12	J4016-18	1	Solid	20.03	20	23.1	8.93	07/17/2018	11:22
13	J4016-21	1	Solid	20.02	20	22.9	9.62	07/17/2018	11:24
14	J4016-24	1	Solid	20.03	20	23.1	9.01	07/17/2018	11:26
15	CCV2	1	Water			21.8	12.00	07/17/2018	11:28

1896687

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY071718

WorkList ID : 114557

Date : 7/17/2018 10:45:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/16/2018	Solid	J4016-03	Corrosivity	Cool 4 deg C	DIST01	L42	SB2-U-23-13	07/16/2018	9045C
07/16/2018	Solid	J4016-06	Corrosivity	Cool 4 deg C	DIST01	L42	SB2-L-12-7	07/16/2018	9045C
07/16/2018	Solid	J4016-09	Corrosivity	Cool 4 deg C	DIST01		SB3-U-23-13	07/16/2018	9045C
07/16/2018	Solid	J4016-12	Corrosivity	Cool 4 deg C	DIST01	L42	SB3-L-12-7	07/16/2018	9045C
07/16/2018	Solid	J4016-15	Corrosivity	Cool 4 deg C	DIST01	L42	SB4-U-23-13	07/16/2018	9045C
07/16/2018	Solid	J4016-18	Corrosivity	Cool 4 deg C	DIST01	L42	SB4-L-12-7	07/16/2018	9045C
07/16/2018	Solid	J4016-21	Corrosivity	Cool 4 deg C	DIST01	L42	SB5-U-23-13	07/16/2018	9045C
07/16/2018	Solid	J4016-24	Corrosivity	Cool 4 deg C	DIST01	L42	SB5-L-12-7	07/16/2018	9045C

Date/Time 07/17/18 10:50
 Received by: SP
 Relinquished by: CP

Date/Time 07/17/18 11:35
 Received by: CP
 Relinquished by: SP