

LB96690

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/17/2018 12:59

Test: Total CN ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.080	0.0	0.094	
ICB1	0.444	0.0	0.004	
CCV1	245.172	0.0	0.232	
CCB1	0.586	0.0	0.004	
PB111192BL	0.811	0.0	0.004	
PB111192BS	97.470	0.0	0.094	
LOWPB111192	10.149	0.0	0.013	
HIGHPB111192	494.743	0.0	0.464	
J4006-01	40.968	0.0	0.042	
J4006-01DUP	40.447	0.0	0.041	
J4006-01MS	82.298	0.0	0.080	
J4006-01MSD	85.049	0.0	0.083	
J4008-01	17.218	0.0	0.020	
CCV2	261.110	0.0	0.246	
CCB2	1.016	0.0	0.004	

N	15
Mean	98.304
SD	137.5164
CV%	139.89

Aquakem v. 7.2AQ1

Results from time period:

Tue Jul 17 10:34:52 2018

Tue Jul 17 12:59:15 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.0195	µg/l	7/17/2018 11:09:59	
5.0PPBCN	A	Total CN	P	5.2986	µg/l	7/17/2018 11:10:00	
10PPBCN	A	Total CN	P	9.9917	µg/l	7/17/2018 11:10:01	
50PPBCN	A	Total CN	P	49.615	µg/l	7/17/2018 11:10:02	
100PPBCN	A	Total CN	P	99.3708	µg/l	7/17/2018 11:10:03	
250PPBCN	A	Total CN	P	251.1636	µg/l	7/17/2018 11:10:04	
500PPBCN	A	Total CN	P	499.5797	µg/l	7/17/2018 11:10:05	
ICV1	S	Total CN	P	97.08	µg/l	7/17/2018 12:54:53	
ICB1	S	Total CN	P	0.4437	µg/l	7/17/2018 12:54:54	
CCV1	S	Total CN	P	245.1723	µg/l	7/17/2018 12:54:55	
CCB1	S	Total CN	P	0.5861	µg/l	7/17/2018 12:54:56	
PB111192BL	S	Total CN	P	0.8108	µg/l	7/17/2018 12:54:57	
PB111192BS	S	Total CN	P	97.4698	µg/l	7/17/2018 12:54:58	
LOWPB111192	S	Total CN	P	10.1494	µg/l	7/17/2018 12:54:59	
HIGHPB111192	S	Total CN	P	494.7433	µg/l	7/17/2018 12:55:00	
J4006-01	S	Total CN	P	40.968	µg/l	7/17/2018 12:55:01	
J4006-01DUP	S	Total CN	P	40.4475	µg/l	7/17/2018 12:55:02	
J4006-01MS	S	Total CN	P	82.2975	µg/l	7/17/2018 12:55:03	
J4006-01MSD	S	Total CN	P	85.0492	µg/l	7/17/2018 12:59:12	
J4008-01	S	Total CN	P	17.2183	µg/l	7/17/2018 12:59:13	
CCV2	S	Total CN	P	261.1096	µg/l	7/17/2018 12:59:14	
CCB2	S	Total CN	P	1.0155	µg/l	7/17/2018 12:59:15	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/17/2018 11:21


Test Total CN

Accepted 7/17/2018 11:10

Factor 1075
Bias 0.004

Coeff. of det. 0.999989

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0195	0.0000	
2	5.0PPBCN	0.008	5.2986	5.0000	
3	10PPBCN	0.013	9.9917	10.0000	
4	50PPBCN	0.050	49.6150	50.0000	
5	100PPBCN	0.096	99.3708	100.0000	
6	250PPBCN	0.237	251.1636	250.0000	
7	500PPBCN	0.468	499.5797	500.0000	