



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96692

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	81.2	07/17/2018	13:20
2	J4006-02		1	8	>100.0	>212.0	>212.0	07/17/2018	13:50
3	J4006-02DUP		1	8	>100.0	>212.0	>212.0	07/17/2018	14:15
4	J4008-02		1	7	>100.0	>212.0	>212.0	07/17/2018	14:42

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96692

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP71718

WorkList ID : 114566

Date : 7/17/2018 12:43:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/19/2018	Water	J4006-02 C	Flash Point	Cool 4 deg C	PSEG01	K62	RR-BASELINE-WATER	07/16/2018	1010A
07/19/2018	Water	J4008-02 C	Flash Point	Cool 4 deg C	PSEG01	K61	KG-FM1026	07/16/2018	1010A

Date/Time

Received by:

Relinquished by:

7/17/18 12:45 PM

KS

CP

Date/Time

Received by:

Relinquished by:

7/17/18 14:45

CP

KS