


LB 966 95

245.1

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 2.7260e-004

C= -4.4524e-003

Rho= 0.9998933

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	-50	0.000	-50				
0.05	0.050	0.047	-0.003	188	0.0 %	188				
0.2	0.200	0.217	0.017	814	0.0 %	814				
2.5	2.500	2.536	0.036	9319	0.0 %	9319				
5.0	5.000	5.017	0.017	18421	0.0 %	18421				
7.5	7.500	7.383	-0.117	27099	0.0 %	27099				
10.0	10.000	10.068	0.068	36949	0.0 %	36949				

LB96695 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Jul 2018 15:27:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.0000	245.1	PPB	17 Jul 2018 15:37:35	S	Std
0.05 - 1	50.05	188	-	0.0500	245.1	PPB	17 Jul 2018 15:39:44	S	Std
0.2 - 1	50.2	814	-	0.2000	245.1	PPB	17 Jul 2018 15:41:53	S	Std
2.5 - 1	52.5	9319	-	2.5000	245.1	PPB	17 Jul 2018 15:44:00	S	Std
5.0 - 1	55	18421	-	5.0000	245.1	PPB	17 Jul 2018 15:46:08	S	Std
7.5 - 1	57.5	27099	-	7.5000	245.1	PPB	17 Jul 2018 15:50:44	S	Std
10.0 - 1	510	36949	-	10.0000	245.1	PPB	17 Jul 2018 16:05:52	S	Std
ICV55 - 1	ICV55	14403	3.9218	-245.1		PPB	17 Jul 2018 16:13:10	U	SMPL
ICB55 - 1	ICB55	-172	-0.0513	-245.1		PPB	17 Jul 2018 16:15:16	U	SMPL
CCV13 - 1	CCV13	18766	5.1112	-245.1		PPB	17 Jul 2018 16:17:24	U	SMPL
CCB13 - 1	CCB13	-182	-0.0541	-245.1		PPB	17 Jul 2018 16:19:30	U	SMPL
CRA - 1	CRA	200	0.0501	-245.1		PPB	17 Jul 2018 16:21:38	U	SMPL
HighStd - 1	HighStd	37472	10.2105	-245.1		PPB	17 Jul 2018 16:23:44	U	SMPL
ChkStd - 1	ChkStd	25002	6.8112	-245.1		PPB	17 Jul 2018 16:25:51	U	SMPL
PB111176BL - 1	PBW	-198	-0.0584	-245.1		PPB	17 Jul 2018 16:28:00	U	SMPL
PB111176BS - 1	LCSW	13563	3.6929	-245.1		PPB	17 Jul 2018 16:30:09	U	SMPL
J3990-01 - 1	071218-DBRI-F-D-B	-245	-0.0712	-245.1		PPB	17 Jul 2018 16:32:17	U	SMPL
J3990-02 - 1	071218-DBRI-F-D-M	-179	-0.0532	-245.1		PPB	17 Jul 2018 16:34:25	U	SMPL
J3990-03 - 1	071218-DBRI-F-D-S	-169	-0.0505	-245.1		PPB	17 Jul 2018 16:36:33	U	SMPL
J3990-04 - 1	071218-DBRI-F-U-B	-141	-0.0429	-245.1		PPB	17 Jul 2018 16:38:41	U	SMPL
J3990-05 - 1	071218-DBRI-F-U-M	-169	-0.0505	-245.1		PPB	17 Jul 2018 16:40:48	U	SMPL
CCV14 - 1	CCV14	18599	5.0657	-245.1		PPB	17 Jul 2018 16:42:55	U	SMPL
CCB14 - 1	CCB14	-189	-0.0560	-245.1		PPB	17 Jul 2018 16:45:01	U	SMPL
J3990-06 - 1	071218-DBRI-F-U-S	-171	-0.0511	-245.1		PPB	17 Jul 2018 16:47:10	U	SMPL
J3990-07 - 1	071218-TIPC-F-D-B	-166	-0.0497	-245.1		PPB	17 Jul 2018 16:49:16	U	SMPL
J3990-08 - 1	071218-TIPC-F-D-S	-145	-0.0440	-245.1		PPB	17 Jul 2018 16:51:23	U	SMPL
J3990-09 - 1	071218-TIPC-F-U-B	-168	-0.0502	-245.1		PPB	17 Jul 2018 16:53:30	U	SMPL
J3990-10 - 1	071218-TIPC-F-U-S	-171	-0.0511	-245.1		PPB	17 Jul 2018 16:55:37	U	SMPL
J3990-10DUP - 1	071218-TIPC-F-U-SDUP	-164	-0.0492	-245.1		PPB	17 Jul 2018 16:57:44	U	SMPL
J3990-10MS - 1	071218-TIPC-F-U-SMS	15637	4.2582	-245.1		PPB	17 Jul 2018 16:59:52	U	SMPL
J3990-10MSD - 1	071218-TIPC-F-U-SMSD	15357	4.1819	-245.1		PPB	17 Jul 2018 17:02:00	U	SMPL
J3990-10LX5 - 1		-179	-0.0532	-245.1		PPB	17 Jul 2018 17:04:09	U	SMPL
J3990-10A - 1		16513	4.4970	-245.1		PPB	17 Jul 2018 17:06:16	U	SMPL
CCV15 - 1	CCV15	18484	5.0343	-245.1		PPB	17 Jul 2018 17:08:23	U	SMPL
CCB15 - 1	CCB15	-193	-0.0571	-245.1		PPB	17 Jul 2018 17:10:30	U	SMPL