



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96699

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/17/2018	16:10
2	CAL2	1	Water			21.6	7.00	07/17/2018	16:12
3	CAL3	1	Water			21.8	10.03	07/17/2018	16:14
4	ICV	1	Water			21.7	7.00	07/17/2018	16:16
5	CCV1	1	Water			21.8	2.00	07/17/2018	16:18
6	J4034-01	1	Water			23.3	4.46	07/17/2018	16:20
7	J4034-01DUP	1	Water			23.4	4.47	07/17/2018	16:22
8	CCV2	1	Water			21.6	12.00	07/17/2018	16:24

LB 96699

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH071718 WorkList ID : 114576 Date : 7/17/2018 3:48:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/09/2018	Water	J4034-01	A pH	Cool 4 deg C	STER10	L52	CELL-A-POOL-WATER(DE-IC	07/09/2018	SM 4500-PH B

Date/Time 07/17/18 4:00 PM
Received by: SP
Relinquished by: CP

Date/Time 07/17/18 4:30 PM
Received by: CP
Relinquished by: SP