



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96704

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	07/17/2018	16:10
2	CAL2	1	Water			21.6	7.00	07/17/2018	16:12
3	CAL3	1	Water			21.8	10.03	07/17/2018	16:14
4	ICV	1	Water			21.7	7.00	07/17/2018	16:16
5	CCV1	1	Water			21.8	2.00	07/17/2018	17:10
6	J4018-02	1	Water			24.3	6.75	07/17/2018	17:12
7	J4018-02DUP	1	Water			24.4	6.76	07/17/2018	17:14
8	CCV2	1	Water			21.6	12.00	07/17/2018	17:16

WORKLIST(Hardcopy Internal Chain)

LB96704

WorkList Name : pH071718-2

WorkList ID : 114577

Date : 7/17/2018 3:48:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/17/2018	Water	J4018-02	pH	Cool 4 deg C	PSEG01	K62	DRUM-3-4	07/17/2018	9040C

Date/Time 07/17/18 5:00 PM
 Received by: SP
 Relinquished by: CP

Date/Time 07/17/18 5:30
 Received by: CP
 Relinquished by: SP